

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004191

Location: Valley City PD

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
- F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR 3455 Lot No. 14270 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR 5132 Lot No. 1CS#2 Solution 0.10 AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 33913080A4 Cyl No. 19 Exp. Date 01/01/2016  
Test 1 0.080 Test 4 0.080 Test 7 0.080  
Test 2 0.080 Test 5 0.080 Test 8 0.079  
Test 3 0.079 Test 6 0.080 Test 9 0.080  
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 940 psi Reg 950 psi

K. Remarks/Maintenance record: Annual inspection, cleaned breath hose screen.

15 Apr 15  
Date

17 APR 15  
Date

Deb Kashur  
Field Inspector's Signature

Charles E. Edm ✓  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = VALL      8164.13.00 06/09  
04/15/2015      01:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:21
02 Std. Sol.	0.020	01:22
03 Room Air	0.000	01:22
04 Std. Sol.	0.020	01:23
05 Room Air	0.000	01:23
06 Std. Sol.	0.020	01:24
07 Room Air	0.000	01:25

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 02

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = VALL      8164.13.00 06/09  
04/15/2015      01:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:27
02 Std. Sol.	0.304	01:28
03 Room Air	0.000	01:29
04 Std. Sol.	0.304	01:30
05 Room Air	0.000	01:30
06 Std. Sol.	0.301	01:31
07 Room Air	0.000	01:31

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 02

Oper No. = 777777

*Deb Keshur*

Operator Signature  
DEB KASHUR

Remarks:

*High AC ~~0.0~~ 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = VALL                    8164.13.00 06/09  
04/15/2015                            01:34

Test	AC	Time
01 Room Air	0.000	01:35
02 *Subject Test	INT*	01:35
03 Room Air	0.000	01:36

\*Invalid Test  
Interferent Detected

Sub Name = TEST, NA NA  
Sub DOB = 08/13/1980  
Sub Sex = Unknown                    Weight = NA  
Test = DUI                                Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 33913080A4  
Cyl No = 19  
Expiration Date = 01/01/2016  
County = 02                                Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10Ac + 0.05%  
acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = VALL                    8164.13.00 06/09  
04/15/2015                            01:37

Test	AC	Time
01 Diagnostic	OK	01:38
02 Room Air	RFI*	01:38
03 Room Air	0.000	01:38

\*Invalid Test  
  Inhibited - RFI

Sub Name = RFI, NA NA  
Sub DOB = 08/13/1980  
Sub Sex = Unknown                    Weight = NA  
Test = DUI                              Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 33913080A4  
Cyl No = 19  
Expiration Date = 01/01/2016  
County = 02                              Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = VALL                    8164.13.00 06/09  
04/15/2015                            01:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:03
02 Std. Gas	0.080	01:04
03 Room Air	0.000	01:05
04 Std. Gas	0.080	01:05
05 Room Air	0.000	01:06
06 Std. Gas	0.079	01:06
07 Room Air	0.000	01:07

Lot No = 33913080A4  
Cyl No = 19  
Exp Date = 01/01/2016  
County = 02                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = VALL                    8164.13.00 06/09  
04/15/2015                            01:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:07
02 Std. Gas	0.080	01:08
03 Room Air	0.000	01:08
04 Std. Gas	0.080	01:09
05 Room Air	0.000	01:09
06 Std. Gas	0.080	01:10
07 Room Air	0.000	01:10

Lot No = 33913080A4  
Cyl No = 19  
Exp Date = 01/01/2016  
County = 02                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = VALL      8164.13.00 06/09  
04/15/2015      01:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:13
02 Std. Gas	0.080	01:14
03 Room Air	0.000	01:14
04 Std. Gas	0.079	01:15
05 Room Air	0.000	01:15
06 Std. Gas	0.080	01:16
07 Room Air	0.000	01:16

Lot No = 33913080A4  
Cyl No = 19  
Exp Date = 01/01/2016  
County = 02      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*