

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005361

Location: JAME - upstairs in jail

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5132 Lot No. 105#2 Solution 0.10 AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 33913080A4 Cyl No. 17 Exp. Date 01/01/2016
Test 1 0.079 Test 4 0.080 Test 7 0.080
Test 2 0.079 Test 5 0.079 Test 8 0.080
Test 3 0.080 Test 6 0.079 Test 9 0.080
Average 0.080

J. Tank pressure, Level 3 Func D Sub G DK 844
~~P-1022~~ psi Reg 850 psi

K. Remarks/Maintenance record: Annual inspection, cleaned breath hose screen, replaced breath hose o-ring

15 Apr 15
Date

17 APR 15 ✓
Date

Deb Kashur
Field Inspector's Signature

Charles E. [Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 17
Standard Expiration? 01/01/2016
Oper No? 777777

Flow Cal. Date: 08/19/2011
Slope 645
Intercept -487053

IR Calibration Date: 01/13/2012
 3um 9um

0th Coef(*100): -5818 -10599
1st Coef(*100): 251241 123805
2nd Coef(*100): 7259 2455
H2O adj(mg/l*10k): 715 798

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:30
02 Std. Sol.	0.022	11:31
03 Room Air	0.000	11:31
04 Std. Sol.	0.021	11:32
05 Room Air	0.000	11:33
06 Std. Sol.	0.022	11:33
07 Room Air	0.000	11:34

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 201501B
County = 47 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020AC

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Sol.	0.309	11:42
03 Room Air	0.000	11:42
04 Std. Sol.	0.305	11:43
05 Room Air	0.000	11:44
06 Std. Sol.	0.304	11:44
07 Room Air	0.000	11:45

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 47

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:47

Test	AC	Time
01 Room Air	0.000	11:48
02 *Subject Test	INT*	11:48
03 Room Air	0.000	11:49

*Invalid Test
Interferent Detected

Sub Name = TEST, NA NA
Sub DOB = 08/12/1980
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 33913080A4
Cyl No = 17
Expiration Date = 01/01/2016
County = 47 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:49

Test	AC	Time
01 Diagnostic	OK	11:50
02 Room Air	RFI*	11:51
03 Room Air	0.000	11:51

*Invalid Test
 Inhibited - RFI

Sub Name = RFI, NA NA
Sub DOB = 08/13/1980
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 33913080A4
Cyl No = 17
Expiration Date = 01/01/2016
County = 47 Oper No. = 77777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:23
02 Std. Gas	0.080	11:23
03 Room Air	0.000	11:23
04 Std. Gas	0.079	11:24
05 Room Air	0.000	11:24
06 Std. Gas	0.079	11:25
07 Room Air	0.000	11:25

Lot No = 33913080A4
Cyl No = 17
Exp Date = 01/01/2016
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/15/2015 11:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Gas	0.080	11:26
03 Room Air	0.000	11:27
04 Std. Gas	0.080	11:27
05 Room Air	0.000	11:28
06 Std. Gas	0.080	11:28
07 Room Air	0.000	11:29

Lot No = 33913080A4
Cyl No = 17
Exp Date = 01/01/2016
County = 47 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC