

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005950

Location: SHER

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5132 Lot No. ICS#2 Solution 0.10AC + 0.05% ACE ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. <u>14214080A6</u>	Cyl No. <u>60</u>	Exp. Date <u>7/5/16</u>
Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.082</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

J. Tank pressure, Level 3 Func D Sub G P: 1010 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

4.14.15  
Date  
17 Apr 15 ✓  
Date

Charles E. Egan  
Field Inspector's Signature  
Deb Kashner  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = SHER                    8164.13.00 06/09  
04/14/2015                            15:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:29
02 Std. Sol.	0.019	15:30
03 Room Air	0.000	15:31
04 Std. Sol.	0.019	15:31
05 Room Air	0.000	15:32
06 Std. Sol.	0.019	15:33
07 Room Air	0.000	15:33


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 42

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = SHER      8164.13.00 06/09  
04/14/2015      15:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:35
02 Std. Sol.	0.298	15:35
03 Room Air	0.000	15:36
04 Std. Sol.	0.300	15:37
05 Room Air	0.000	15:37
06 Std. Sol.	0.302	15:38
07 Room Air	0.000	15:38


08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 42

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = SHER                    8164.13.00 06/09  
04/14/2015                            15:26

Test	AC	Time
01 Room Air	0.000	15:27
02 *Subject Test	INT*	15:27
03 Room Air	0.000	15:28

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 14214080A6  
Cyl No = 60  
Expiration Date = 07/05/2016  
County = 42                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10 AC + 0.05% ACETONE*  
Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = SHER                    8164.13.00 06/09  
04/14/2015                            15:23

Test	AC	Time
01 Diagnostic	OK	15:24
02 Room Air	RFI*	15:24
03 Room Air	0.000	15:25

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 14214080A6  
Cyl No = 60  
Expiration Date = 07/05/2016  
County = 42                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = SHER      8164.13.00 06/09  
04/14/2015      15:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:20
02 Std. Gas	0.081	15:21
03 Room Air	0.000	15:21
04 Std. Gas	0.081	15:22
05 Room Air	0.000	15:22
06 Std. Gas	0.081	15:23
07 Room Air	0.000	15:23


Lot No = 14214080A6

Cyl No = 60

Exp Date = 07/05/2016

County = 42

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = SHER      8164.13.00 06/09  
04/14/2015      15:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:43
02 Std. Gas	0.082	15:44
03 Room Air	0.000	15:44
04 Std. Gas	0.082	15:45
05 Room Air	0.000	15:45
06 Std. Gas	0.082	15:46
07 Room Air	0.000	15:46


Lot No = 14214080A6

Cyl No = 60

Exp Date = 07/05/2016

County = 42

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = SHER      8164.13.00 06/09  
04/14/2015      15:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:47
02 Std. Gas	0.081	15:47
03 Room Air	0.000	15:48
04 Std. Gas	0.081	15:48
05 Room Air	0.000	15:48
06 Std. Gas	0.081	15:49
07 Room Air	0.000	15:49

Lot No = 14214080A6  
Cyl No = 60  
Exp Date = 07/05/2016  
County = 42      Oper No. = 666666

  
Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080Ac*