

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003072 Location: TOXL

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3144</u> Lot No. <u>—</u> Rep AC <u>0.021</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>—</u> Rep AC <u>0.294</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3452</u> Lot No. <u>—</u> Rep AC <u>0.10 AC + Acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>15410080A1</u>	Cyl No. <u>19</u>	Exp. Date <u>7-1-12</u>
Test 1 <u>0.083</u>	Test 4 <u>0.083</u>	Test 7 <u>0.081</u>
Test 2 <u>0.083</u>	Test 5 <u>0.082</u>	Test 8 <u>0.082</u>
Test 3 <u>0.083</u>	Test 6 <u>0.082</u>	Test 9 <u>0.083</u>
Average <u>0.082</u>		

L. Tank pressure, Level 3 Func D Sub G P: 262 psi Reg 262 psi

M. Remarks/Maintenance record: Annual Inspection

4-18-11
Date

Deb Kashner
Field Inspector's Signature

18 Apr 11
Date

Deb Kashner
Reviewed by

80-003072
P. 125 of 133
JFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:20

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 19
Standard Expiration? 07/01/2012
Oper No? 555555

Flow Cal. Date: 03/09/2011
Slope 682
Intercept -620512

IR Calibration Date: 03/09/2011
 3um 9um

0th Coef(*100): -24766 -17499
1st Coef(*100): 264559 138630
2nd Coef(*100): 1876 1114
H2O adj(mg/l*10k): 497 331

***** Printer Test End *****



Operator Signature
ASHLEY AYD

Remarks: Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:22
02 Std. Sol.	0.019	13:23
03 Room Air	0.000	13:23
04 Std. Sol.	0.020	13:24
05 Room Air	0.000	13:25
06 Std. Sol.	0.019	13:25
07 Room Air	0.000	13:26

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Low AC 0.021 AC*

Form 106-I8000

*80-003072
P. 127 of 133
FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Sol.	0.295	13:28
03 Room Air	0.000	13:29
04 Std. Sol.	0.298	13:29
05 Room Air	0.000	13:30
06 Std. Sol.	0.300	13:30
07 Room Air	0.000	13:31


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Linearity test 0.294AC*

Form 106-I8000

*80-003072
p. 128 of 133
FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:32

Test	AC	Time
01 Room Air	0.000	13:33
02 *Subject Test	INT*	13:34
03 Room Air	0.000	13:34

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 19
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: *Interferent check 0.10 AC + Acetone*

Form 106-I8000

80-003072
P-129 of 133
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:35

Test	AC	Time
01 Diagnostic	OK	13:36
02 Room Air	RFI*	13:37
03 Room Air	0.000	13:37

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 15410080A1
Cyl No = 19
Expiration Date = 07/01/2012
County = 08 Oper No. = 555555

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ASHLEY AYD

Remarks: RFI

Form 106-I8000

80-003072
p.130 of 133
YAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Gas	0.083	13:39
03 Room Air	0.000	13:39
04 Std. Gas	0.083	13:40
05 Room Air	0.000	13:40
06 Std. Gas	0.083	13:40
07 Room Air	0.000	13:41

Lot No = 15410080A1
Cyl No = 19
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration Check 0.080 AC

Form 106-I8000

80-003072
p. 131 of 133
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Gas	0.083	13:44
03 Room Air	0.000	13:45
04 Std. Gas	0.082	13:45
05 Room Air	0.000	13:46
06 Std. Gas	0.082	13:46
07 Room Air	0.000	13:47

Lot No = 15410080A1
Cyl No = 19
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration Check 0.080AC

Form 106-I8000

80-003072
p. 132 of 133
FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
04/18/2011 13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:49
02 Std. Gas	0.081	13:49
03 Room Air	0.000	13:50
04 Std. Gas	0.082	13:50
05 Room Air	0.000	13:51
06 Std. Gas	0.083	13:51
07 Room Air	0.000	13:52

Lot No = 15410080A1
Cyl No = 19
Exp Date = 07/01/2012
County = 08 Oper No. = 555555



Operator Signature
ASHLEY AYD

Remarks: Calibration Check 0.080AC

Form 106-I8000

80-003072
P. 133 of 133
FM