

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004938

Location: ToxL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 10270 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 11110 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3377 Lot No. Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080A1 Cyl No. 65 Exp. Date 10/01/2013
Test 1 0.080 Test 4 6.079 Test 7 0.080
Test 2 0.080 Test 5 0.080 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 6.079
Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 570 psi Reg 575 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

19 APR 12
Date

Charles E. Eden
Field Inspector's Signature

19 Apr 2012
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 10:50

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 65
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 01/25/2010
Slope 712
Intercept -604165

IR Calibration Date: 11/10/2010
 3um 9um

0th Coef(*100): -23938 -24123
1st Coef(*100): 266075 133792
2nd Coef(*100): 1429 981
H2O adj(mg/l*10k): 699 681

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:06
02 Std. Sol.	0.019	11:24
03 Room Air	0.000	11:24
04 Std. Sol.	0.019	11:25
05 Room Air	0.000	11:25
06 Std. Sol.	0.019	11:26
07 Room Air	0.000	11:27


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Sol.	0.290	11:30
03 Room Air	0.000	11:31
04 Std. Sol.	0.290	11:31
05 Room Air	0.000	11:32
06 Std. Sol.	0.290	11:33
07 Room Air	0.000	11:33


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:40

Test	AC	Time
01 Room Air	0.000	11:41
02 *Subject Test	INT*	11:42
03 Room Air	0.000	11:42

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 1234567890
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT CHECK
0.100AC + ACETONE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:43

Test	AC	Time
01 Diagnostic	OK	11:45
02 Room Air	RFI*	11:45
03 Room Air	0.000	11:45

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 1234567890
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Gas	0.080	11:49
03 Room Air	0.000	11:49
04 Std. Gas	0.080	11:50
05 Room Air	0.000	11:50
06 Std. Gas	0.079	11:51
07 Room Air	0.000	11:51

Lot No = 23411080A1
Cyl No = 65
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:54
02 Std. Gas	0.079	11:55
03 Room Air	0.000	11:55
04 Std. Gas	0.080	11:56
05 Room Air	0.000	11:56
06 Std. Gas	0.079	11:56
07 Room Air	0.000	11:57

Lot No = 23411080A1
Cyl No = 65
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
 6.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = TOXL 8164.13.00 06/09
04/19/2012 11:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:59
02 Std. Gas	0.080	11:59
03 Room Air	0.000	12:00
04 Std. Gas	0.079	12:00
05 Room Air	0.000	12:01
06 Std. Gas	0.079	12:01
07 Room Air	0.000	12:01

Lot No = 23411080A1
Cyl No = 65
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC