

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004944

Location: TOTAL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>55</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.082</u>
Test 3 <u>0.080</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 720 psi Reg 725 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

17 APR 12
Date

Charles E. Edin
Field Inspector's Signature

18 Apr 2012
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 16:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 55
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 07/01/2011
Slope 685
Intercept -634740

IR Calibration Date: 07/01/2011
 3um 9um

0th Coef(*100): -22591 -17234
1st Coef(*100): 265915 132691
2nd Coef(*100): 2841 1562
H2O adj(mg/l*10k): 559 425

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 16:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:37
02 Std. Sol.	0.019	16:38
03 Room Air	0.000	16:38
04 Std. Sol.	0.020	16:39
05 Room Air	0.000	16:39
06 Std. Sol.	0.021	16:40
07 Room Air	0.000	16:41

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 16:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:49
02 Std. Sol.	0.310	16:50
03 Room Air	0.000	16:50
04 Std. Sol.	0.310	16:51
05 Room Air	0.000	16:51
06 Std. Sol.	0.310	16:52
07 Room Air	0.000	16:53

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 16:55

Test	AC	Time
01 Room Air	0.000	16:56
02 *Subject Test	INT*	16:57
03 Room Air	0.000	16:57

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 123456
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK
0.100 AC + ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 16:58

Test	AC	Time
01 Diagnostic	OK	16:59
02 Room Air	RFI*	17:00
03 Room Air	0.000	17:00

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 123456
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 17:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:04
02 Std. Gas	0.080	17:04
03 Room Air	0.000	17:05
04 Std. Gas	0.081	17:05
05 Room Air	0.000	17:05
06 Std. Gas	0.080	17:06
07 Room Air	0.000	17:06

Lot No = 23411080A1
Cyl No = 55
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 17:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:07
02 Std. Gas	0.081	17:07
03 Room Air	0.000	17:08
04 Std. Gas	0.080	17:08
05 Room Air	0.000	17:09
06 Std. Gas	0.082	17:09
07 Room Air	0.000	17:10


Lot No = 23411080A1

Cyl No = 55

Exp Date = 10/01/2013

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = TOXL 8164.13.00 06/09
04/17/2012 17:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:12
02 Std. Gas	0.081	17:12
03 Room Air	0.000	17:13
04 Std. Gas	0.082	17:13
05 Room Air	0.000	17:14
06 Std. Gas	0.081	17:14
07 Room Air	0.000	17:14

Lot No = 23411080A1
Cyl No = 55
Exp Date = 10/01/2013
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC