



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (1-2010)

Serial Number: <u>80-004947</u>	Instrument Location: <u>Mohl</u>
Reason for Install/Repair:	
<input checked="" type="checkbox"/> Install After Lab Repair/Inspection	
<input type="checkbox"/> Other (Specify) _____	

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: \_\_\_\_\_

*Bryce Venable*  
Field Inspector Signature

~~4-14-11~~ 4-19-11 DK  
Date

*Deb Keshur*  
Reviewed By

19 Apr 2011  
Date

80-004947  
P. 72 of 76  
JRM




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004947  
Location = MOHL                    8164.13.00 06/09  
04/14/2011                            16:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:59
02 Std. Gas	0.081	16:59
03 Room Air	0.000	17:00
04 Std. Gas	0.080	17:00
05 Room Air	0.000	17:01
06 Std. Gas	0.080	17:01
07 Room Air	0.000	17:02

Lot No = 659358  
Cyl No = 27  
Exp Date = 09/03/2011  
County = 38                            Oper No. = 039909

  
\_\_\_\_\_  
Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004947  
P. 74 of 76  
VH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004947  
Location = MOHL                8164.13.00 06/09  
04/14/2011                        17:03

Test	AC	Time
01 Diagnostic	OK	17:06
02 Room Air	RFI*	17:06
03 Room Air	0.000	17:07

\*Invalid Test  
Inhibited - RFI

Sub Name = DOE, JOHN A  
Sub DOB = 08/31/1938  
Sub Sex = Male                    Weight = 122  
Test = DUI                        Cit = 000000  
Dr. Lic. = ND/NA  
Lot No = 659358  
Cyl No = 27  
Expiration Date = 09/03/2011  
County = 38                        Oper No. = 039909

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004947  
p. 75 of 76  
FM

19 April 2011

I called Barry Vannatta to inform him that he didn't send in the Print Test with his other installation paperwork. He remembers thinking he sent it in, but also remembers throwing one away. Officer Vannatta did a Print Test as we spoke and faxed it to me. I informed him that I would have to change the installation date on the Form 104-G.

*Deb Keshner*

80-004947  
P. 76 of 76  
KM