

**INTOXILYZER 8000 INSPECTION**

Serial No.: 80-004934 Location: TOXL

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>10290</u> Rep AC <u>0.020</u>                                     | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3803</u> Lot No. <u>11110</u> Rep AC <u>0.300</u>                            | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>—</u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>23411080A1</u>	Cyl No. <u>49</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.082</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 821 psi Reg 820 psi

M. Remarks/Maintenance record: Annual inspection

19 Apr 2012  
Date  
19 APR 12  
Date

Deb Keshner  
Field Inspector's Signature  
Charles E. Ed...  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-004934  
Location = TOXL                8164.13.00 06/09  
04/19/2012                      11:06

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?         No  
Display Volume?             No  
Disable On Memfull?         Yes  
# of Print Copies?          1  
Select Std (D/W/I)?         Dry  
Standard Value?             0.080  
Standard Lot #?             23411080A1  
Standard Cyl #?             49  
Standard Expiration?        10/01/2013  
Oper No?                    777777

Flow Cal. Date:             01/20/2010  
Slope                        698  
Intercept                   -694859

IR Calibration Date:        07/07/2011  
                              3um                    9um

-----  
0th Coef(\*100):       -8226                   -6096  
1st Coef(\*100):       244617                   123046  
2nd Coef(\*100):       5840                    2321  
H2O adj(mg/l\*10k):   663                     679

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = TOXL      8164.13.00 06/09  
04/19/2012      11:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:24
02 Std. Sol.	0.021	11:25
03 Room Air	0.000	11:26
04 Std. Sol.	0.021	11:27
05 Room Air	0.000	11:27
06 Std. Sol.	0.022	11:28
07 Room Air	0.000	11:28

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = TOXL      8164.13.00 06/09  
04/19/2012      11:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:13
02 Std. Sol.	0.303	11:14
03 Room Air	0.000	11:14
04 Std. Sol.	0.303	11:15
05 Room Air	0.000	11:15
06 Std. Sol.	0.304	11:16
07 Room Air	0.000	11:17

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = TOXL      8164.13.00 06/09  
04/19/2012      11:46

Test	AC	Time
01 Room Air	0.000	11:47
02 *Subject Test	INT*	11:47
03 Room Air	0.000	11:48

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 49  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = TOXL      8164.13.00 06/09  
04/19/2012      10:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Gas	0.081	11:00
03 Room Air	0.000	11:00
04 Std. Gas	0.081	11:00
05 Room Air	0.000	11:01
06 Std. Gas	0.082	11:01
07 Room Air	0.000	11:02

Lot No = 23411080A1  
Cyl No = 49  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check      0.080 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = TOXL      8164.13.00 06/09  
04/19/2012      11:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:34
02 Std. Gas	0.081	11:34
03 Room Air	0.000	11:35
04 Std. Gas	0.081	11:35
05 Room Air	0.000	11:36
06 Std. Gas	0.081	11:36
07 Room Air	0.000	11:37

Lot No = 23411080A1  
Cyl No = 49  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = TOXL      8164.13.00 06/09  
04/19/2012      11:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:40
02 Std. Gas	0.082	11:40
03 Room Air	0.000	11:40
04 Std. Gas	0.081	11:41
05 Room Air	0.000	11:41
06 Std. Gas	0.082	11:42
07 Room Air	0.000	11:42

Lot No = 23411080A1  
Cyl No = 49  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000