



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (5-2011)

Serial Number <i>80-003058</i>	Instrument Location <i>Mercer Co 50</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A.  Print Test (Level 1, Function P).
  - B.  ACA Test (Level 1, Function C).
  - C.  RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): *Installation*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Deb Kasher</i>	Date <i>19 Apr 2013</i>
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Edr</i>	Date <i>22 APR 2013</i>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Charles E. Edr* Certified Date: *04 Nov. 16*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = STAT      8164.13.00 06/09  
04/19/2013      12:12

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      16512080A2  
Standard Cyl #?      52  
Standard Expiration?      08/01/2014  
Oper No?      777777

Flow Cal. Date:      04/07/2008  
Slope      698  
Intercept      -678970

IR Calibration Date:      04/10/2013  
   3um      9um


-----  
0th Coef(\*100):      -8620      -6643  
1st Coef(\*100):      266250      133559  
2nd Coef(\*100):      2825      1008  
H2O adj(mg/l\*10k):      499      374

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:



Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = STAT      8164.13.00 06/09  
04/19/2013      12:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:41
02 Std. Gas	0.079	12:42
03 Room Air	0.000	12:42
04 Std. Gas	0.079	12:43
05 Room Air	0.000	12:43
06 Std. Gas	0.080	12:43
07 Room Air	0.000	12:44

Lot No = 16512080A2  
Cyl No = 52  
Exp Date = 08/01/2014  
County = 29      Oper No. = 777777

  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = STAT      8164.13.00 06/09  
04/19/2013      12:39

Test	AC	Time
01 Diagnostic	OK	12:40
02 Room Air	RFI*	12:40
03 Room Air	0.000	12:41

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 16512080A2  
Cyl No = 52  
Expiration Date = 08/01/2014  
County = 29      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000