

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-004200

Location: Tox L

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 523 psi Regulator: 525 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3060 Lot #: 202201A Exp. Date: 1/11/24
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.300 AC
Sim. Ser #: MP3064 Lot #: 202012A Exp. Date: 12/6/22

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3847 Lot #: ICS 6 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 4 Exp. Date: 3/5/23

Test 1 <u>0.084</u>	Test 4 <u>0.084</u>	Test 7 <u>0.083</u>
Test 2 <u>0.084</u>	Test 5 <u>0.083</u>	Test 8 <u>0.084</u>
Test 3 <u>0.083</u>	Test 6 <u>0.083</u>	Test 9 <u>0.084</u>
Average <u>0.084</u>		

C. Remarks/Maintenance: Annual Inspection.

Replaced breath tube o-ring. Replaced simulator return o-ring.
Replaced external battery.

Instrument is acceptable to be used in the field.

Janelle Portschiller
Breath Analyst Signature

21 Apr 22
Date

[Signature]
Reviewed by

04/21/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 09:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 37
Standard Expiration? 04/05/2022
Oper No? 132128

Flow Cal. Date: 06/25/2019
Slope 614
Intercept -551153

IR Calibration Date: 06/25/2019
 3um 9um

0th Coef(*100): -18792 -22164
1st Coef(*100): 267001 140777
2nd Coef(*100): 1903 1083
H2O adj(mg/l*10k): 403 445

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 10:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:21
02 Std. Sol.	0.023	10:22
03 Room Air	0.000	10:23
04 Std. Sol.	0.022	10:23
05 Room Air	0.000	10:24
06 Std. Sol.	0.022	10:25
07 Room Air	0.000	10:25

08 Sim Temp = 34.0°C

Simul Ser No = MP3060

Std Sol No = 202201A

County = 08

Oper No. = 132128

Janelle Ports Cheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low AC Check 0.020*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 10:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Sol.	0.296	10:27
03 Room Air	0.000	10:28
04 Std. Sol.	0.299	10:28
05 Room Air	0.000	10:29
06 Std. Sol.	0.301	10:30
07 Room Air	0.000	10:30

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks:

High AC Check 0.300 ~~0.400~~


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 10:33

Test	AC	Time
01 Room Air	0.000	10:34
02 *Subject Test	INT*	10:34
03 Room Air	0.000	10:35

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 10:38

Test	AC	Time
01 Diagnostic	OK	10:39
02 Room Air	0.000	10:39
03 *Subject Test	RFI*	10:39
04 Room Air	0.000	10:40

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CK
Dr. Lic. = ND/DIS921456
Lot No = 02621080A1
Cyl No = 4
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 11:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:12
02 Std. Gas	0.084	11:13
03 Room Air	0.000	11:13
04 Std. Gas	0.084	11:14
05 Room Air	0.000	11:14
06 Std. Gas	0.083	11:14
07 Room Air	0.000	11:15

Lot No = 02621080A1

Cyl No = 4

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks:

Dry calibration check 1-3

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 11:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.084	11:29
03 Room Air	0.000	11:29
04 Std. Gas	0.083	11:30
05 Room Air	0.000	11:30
06 Std. Gas	0.083	11:31
07 Room Air	0.000	11:31

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128


Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check* *4-300*
4-6


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = TOXL 8164.14.00 09/16
04/21/2022 11:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:37
02 Std. Gas	0.083	11:37
03 Room Air	0.000	11:38
04 Std. Gas	0.084	11:38
05 Room Air	0.000	11:39
06 Std. Gas	0.084	11:39
07 Room Air	0.000	11:40

Lot No = 02621080A1
Cyl No = 4
Exp Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check 7-9*