



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (1-2010)

Serial Number: 80-004954 Instrument Location: FYPD

Reason for Install/Repair:

- Install After Lab Repair/Inspection
 Other (Specify) New Instrument

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
 B. Telephone Line Connected to Intoxilyzer® 8000.
 C. Breath Tube Heated.
 D. Review/Enter Preliminary Data Entry (Level 2, Function E).
 E. Review/Enter Gas Setup (Level 1, Function S).
 F. Print Test (Level 1, Function P). (Attach Test Record.)
 G. ACA Test (Level 1, Function C). (Attach Test Record.)
 H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
I. Repair and/or Maintenance Performed: _____

[Signature]
Field Inspector Signature

04-22-10
Date

Deb Kasher
Reviewed By

10 May 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
04/22/2010 19:49

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 26
Standard Expiration? 12/23/2011
Oper No? 010606

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -671831

IR Calibration Date: 01/19/2010

	3um	9um
0th Coef(*100):	-34134	-33955
1st Coef(*100):	257873	132608
2nd Coef(*100):	1059	850
H2O adj(mg/l*10k):	762	516

***** Printer Test End *****



Operator Signature
CHAD L. HARMON

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
04/22/2010 19:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:51
02 Std. Gas	0.079	19:52
03 Room Air	0.000	19:52
04 Std. Gas	0.081	19:52
05 Room Air	0.000	19:53
06 Std. Gas	0.080	19:53
07 Room Air	0.000	19:54

Lot No = 709352
Cyl No = 26
Exp Date = 12/23/2011
County = 61 Oper No. = 010606



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. 45/46 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
04/22/2010 19:58

Test	AC	Time
01 Diagnostic	OK	19:59
02 Room Air	RFI*	19:59
03 Room Air	RFI*	19:59

*Invalid Test
Inhibited - RFI

Sub Name = YELLOWFAT, JUSTIN WADE
Sub DOB = 11/17/1972
Sub Sex = Male Weight = 210
Test = OTH Cit = NA
Dr. Lic. = ND/YEL722120
Lot No = 709352
Cyl No = 26
Expiration Date = 12/23/2011
County = 61 Oper No. = 010606

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. 46/46 van