



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Richard D. Johnson</u>		
Location: <u>McLean LEC WASH</u>		
Intoxilyzer® Serial Number: <u>80-004196</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>47</u>	Expiration Date: <u>9/30/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) <u>DK 25 Apr 11</u> Test 1: <u>0.081</u> % Test 2: <u>0.081</u> % Test 3: <u>0.081</u> %		

(ATTACH TEST RECORD)

Richard D. Johnson
Chemical Test Operator Signature

4-22-11
Date

Deb Kashner
Reviewed By

25 Apr 2011
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
04/22/2011 10:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:59
02 Std. Gas	0.081	10:59
03 Room Air	0.000	10:59
04 Std. Gas	0.081	11:00
05 Room Air	0.000	11:00
06 Std. Gas	0.081	11:01
07 Room Air	0.000	11:01

Lot No = 659357
Cyl No = 47
Exp Date = 09/03/2011
County = 28 Oper No. = 118707


Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196
P.109 of 104
JRM