

INTOXILYZER 8000 INSPECTION

Serial No.: 80-005361

Location: Jamestown (JAME) adult corrections

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 12210 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 12050 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5191 Lot No. KS#1 Rep AC 0.100 + acetone ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)
23411080A1
Lot No. ~~DK-0.080~~ Cyl No. 42 Exp. Date 10/01/2013
Test 1 0.080 Test 4 0.080 Test 7 0.079
Test 2 0.080 Test 5 0.081 Test 8 0.079
Test 3 0.080 Test 6 0.080 Test 9 0.078
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 769 psi Reg 760 psi

M. Remarks/Maintenance record: Annual inspection, good battery, replaced breath hose o-ring.

18 Apr 2013
Date

Deb Kashner
Field Inspector's Signature

22 APR 2013
Date

Charles E. Egan
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 10:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 42
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 08/19/2011
Slope 645
Intercept -487053

IR Calibration Date: 01/13/2012
 3um 9um

0th Coef(*100): -5818 -10599
1st Coef(*100): 251241 123805
2nd Coef(*100): 7259 2455
H2O adj(mg/l*10k): 715 798

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 11:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:03
02 Std. Sol.	0.019	11:04
03 Room Air	0.000	11:04
04 Std. Sol.	0.020	11:05
05 Room Air	0.000	11:05
06 Std. Sol.	0.020	11:06
07 Room Air	0.000	11:07

08 Sim Temp = 34.0°C

Simul Ser No = DR5132

Std Sol No = 12210

County = 47

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 11:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:10
02 Std. Sol.	0.299	11:11
03 Room Air	0.000	11:12
04 Std. Sol.	0.299	11:13
05 Room Air	0.000	11:13
06 Std. Sol.	0.299	11:14
07 Room Air	0.000	11:14

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 12050

County = 47

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

^{DK}
Calibration Linearity test 0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 10:59

Test	AC	Time
01 Room Air	0.000	10:59
02 *Subject Test	INT*	11:00
03 Room Air	0.000	11:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 42
Expiration Date = 10/01/2013
County = 47 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 10:56

Test	AC	Time
01 Diagnostic	OK	10:57
02 Room Air	RFI*	10:57
03 Room Air	0.000	10:57

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 42
Expiration Date = 10/01/2013
County = 47 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Gas	0.080	10:46
03 Room Air	0.000	10:46
04 Std. Gas	0.080	10:47
05 Room Air	0.000	10:47
06 Std. Gas	0.080	10:47
07 Room Air	0.000	10:48

Lot No = 23411080A1

Cyl No = 42

Exp Date = 10/01/2013

County = 47

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 10:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:50
02 Std. Gas	0.080	10:50
03 Room Air	0.000	10:51
04 Std. Gas	0.081	10:51
05 Room Air	0.000	10:51
06 Std. Gas	0.080	10:52
07 Room Air	0.000	10:52

Lot No = 23411080A1
Cyl No = 42
Exp Date = 10/01/2013
County = 47 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005361
Location = JAME 8164.13.00 06/09
04/18/2013 10:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:53
02 Std. Gas	0.079	10:54
03 Room Air	0.000	10:54
04 Std. Gas	0.079	10:54
05 Room Air	0.000	10:55
06 Std. Gas	0.078	10:55
07 Room Air	0.000	10:56

Lot No = 23411080A1
Cyl No = 42
Exp Date = 10/01/2013
County = 47 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC