

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005943

Location: BSPD

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. ICS2 Rep AC 0.100 AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 40 Exp. Date 08/01/2014

Test 1 <u>0.082</u>	Test 4 <u>0.082</u>	Test 7 <u>0.083</u>
Test 2 <u>0.083</u>	Test 5 <u>0.082</u>	Test 8 <u>0.083</u>
Test 3 <u>0.082</u>	Test 6 <u>0.083</u>	Test 9 <u>0.84^{2k} 0.084</u>

Average 0.083

J. Tank pressure, Level 3 Func D Sub G P: 617 psi Reg 600 psi

K. Remarks/Maintenance record: annual inspection

22 Apr 2014
Date

22 APR 2014
Date

Deb Kashur
Field Inspector's Signature

Charles E. Eden
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:07
02 Std. Sol.	0.021	15:08
03 Room Air	0.000	15:09
04 Std. Sol.	0.022	15:09
05 Room Air	0.000	15:10
06 Std. Sol.	0.022	15:11
07 Room Air	0.000	15:11

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 08

Oper No. = 777777

Deb Keshur

Operator Signature
DEB KASHUR

Remarks:

low AC 0.021AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:18
02 Std. Sol.	0.308	15:19
03 Room Air	0.000	15:20
04 Std. Sol.	0.309	15:21
05 Room Air	0.000	15:21
06 Std. Sol.	0.308	15:22
07 Room Air	0.000	15:23

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:24

Test	AC	Time
01 Room Air	0.000	15:25
02 *Subject Test	INT*	15:25
03 Room Air	0.000	15:26

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 40
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Debs Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:27

Test	AC	Time
01 Diagnostic	OK	15:27
02 Room Air	RFI*	15:28
03 Room Air	0.000	15:28

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 40
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:29
02 Std. Gas	0.082	15:30
03 Room Air	0.000	15:30
04 Std. Gas	0.083	15:31
05 Room Air	0.000	15:31
06 Std. Gas	0.082	15:31
07 Room Air	0.000	15:32

Lot No = 16512080A2
Cyl No = 40
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:33
02 Std. Gas	0.082	15:33
03 Room Air	0.000	15:34
04 Std. Gas	0.082	15:34
05 Room Air	0.000	15:35
06 Std. Gas	0.083	15:35
07 Room Air	0.000	15:36

Lot No = 16512080A2
Cyl No = 40
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005943
Location = BSPD 8164.13.00 06/09
04/22/2014 15:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:37
02 Std. Gas	0.083	15:37
03 Room Air	0.000	15:38
04 Std. Gas	0.083	15:38
05 Room Air	0.000	15:39
06 Std. Gas	0.084	15:39
07 Room Air	0.000	15:39

Lot No = 16512080A2
Cyl No = 40
Exp Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC