



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Ray Weber</u>		
Location: <u>TRAIL CO. S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004200</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>#45</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.080</u> % Test 2: <u>0.080</u> % Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

Ray Weber
 Chemical Test Operator Signature

4/23/2010
 Date

Deb Kashner
 Reviewed By

10 May 2010
 Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = HILL 8164.13.00 06/09
04/23/2010 21:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:24
02 Std. Gas	0.080	21:25
03 Room Air	0.000	21:25
04 Std. Gas	0.080	21:25
05 Room Air	0.000	21:26
06 Std. Gas	0.080	21:26
07 Room Air	0.000	21:27


Lot No = 659358

Cyl No = 45

Exp Date = 09/03/2011

County = 49

Oper No. = 058405



Operator Signature
RAY G. WEBER

Remarks:

Form 106-18000

80-004200
P. 75/75 KM