

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004191

Location: ToxL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3877</u> Lot No. <u>11110</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR377</u> Lot No. <u>—</u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>2341108041</u>	Cyl No. <u>65</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 451 psi Reg 490 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

23 APR 2012
Date

Charles E. Edler
Field Inspector's Signature

23 Apr 2012
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/20/2012 15:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:54
02 Std. Sol.	0.017	15:54
03 Room Air	0.000	15:55
04 Std. Sol.	0.018	15:56
05 Room Air	0.000	15:56
06 Std. Sol.	0.018	15:57
07 Room Air	0.000	15:58

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 0000
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/20/2012 15:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:58
02 Std. Sol.	0.296	15:59
03 Room Air	0.000	16:00
04 Std. Sol.	0.296	16:01
05 Room Air	0.000	16:01
06 Std. Sol.	0.296	16:02
07 Room Air	0.000	16:03

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/23/2012 10:16

Test	AC	Time
01 Room Air	0.000	10:17
02 *Subject Test	INT*	10:17
03 Room Air	0.000	10:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOP
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 65
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100 AC + ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/23/2012 10:20

Test	AC	Time
01 Diagnostic	OK	10:21
02 Room Air	RFI*	10:21
03 Room Air	0.000	10:21

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = ASDFGHJKL;
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 65
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/23/2012 09:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:53
02 Std. Gas	0.079	09:53
03 Room Air	0.000	09:54
04 Std. Gas	0.078	09:54
05 Room Air	0.000	09:55
06 Std. Gas	0.079	09:55
07 Room Air	0.000	09:56


Lot No = 23411080A1

Cyl No = 65

Exp Date = 10/01/2013

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/23/2012 10:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:00
02 Std. Gas	0.079	10:01
03 Room Air	0.000	10:01
04 Std. Gas	0.078	10:02
05 Room Air	0.000	10:02
06 Std. Gas	0.079	10:03
07 Room Air	0.000	10:03

Lot No = 23411080A1
Cyl No = 65
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = TOXL 8164.13.00 06/09
04/23/2012 10:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:08
02 Std. Gas	0.080	10:08
03 Room Air	0.000	10:09
04 Std. Gas	0.080	10:09
05 Room Air	0.000	10:10
06 Std. Gas	0.080	10:10
07 Room Air	0.000	10:11

Lot No = 23411080A1
Cyl No = 65
Exp Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000