CMI INC.

SERVICE WORK ORDER

316 E. 9th St., Owensboro KY 42303
Phone 866-835-0690
Fax: 270-685-6268

RCVD 5-7-14
RCVD VIA FEDX

334008

BILL TO:	SHIP TO: ND Crime Lab
ATTN:	ATTN:
	2641 E. Main Ave
CITY STATE ZIP	CITY Bismarck STATE ND ZIP 58501
CUSTOMER NO. 585013	PHONE 701-328-6159
PHONE	FAX 701-328-6185

MODEL # 8000 | SERIAL # 80-005954 | BILL CODE 3

EXTRA PARTS RCVD: BT, gas box, battery

DESCRIPTION OF PROBLEM: RTC Clock fail during diagnostic

Infrared / Fuel Cell
PRELIMINARY TEST INFORMATION

Date read 14/24/2014
BATTERY-EXT 2.44 V, INT 3.04 V
Reset & following day year read 2023
Consulted engineering & determined system board needs replaced
Gas pressure reading low, half value Res
VERIFIED PROBLEM?

WORK PERFORMED:
Replaced system board for RTC error
Replaced gas pressure sensor

Miscellaneous
PRELIMINARY TEST INFORMATION

WORK PERFORMED:
PDI

PARTS USED

QTY	PART NO.	DESCRIPTION
1	0212706111T	System Board
1	210131	Sensor, Pressure

REPAIR	CAL. ADJ.	CALIBRATON	FINAL
GTW	NO	NO	NO
½ HOURS	HOURS	HOURS	HOURS
8	0.5	1.0	1.0
DATE	DATE	DATE	DATE
May 8 2014	May 9, 14	May 9, 14	May 9, 14

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS

FORM# CMI SWO REV (1)