

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003058

Location: STAT - Stanton

- | | Check When Done |
|---|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) Sim SN <u>DR5115</u> Lot No. <u>12111</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.) Sim SN <u>DR5143</u> Lot No. <u>13080</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) Sim SN <u>DR3846</u> Lot No. <u>IC52</u> Rep AC <u>0.104c + 0.05% acetone</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.) | |

| | | |
|---------------------------|---------------------|-----------------------------|
| Lot No. <u>16512080A2</u> | Cyl No. <u>52</u> | Exp. Date <u>08/01/2014</u> |
| Test 1 <u>0.078</u> | Test 4 <u>0.079</u> | Test 7 <u>0.078</u> |
| Test 2 <u>0.078</u> | Test 5 <u>0.078</u> | Test 8 <u>0.079</u> |
| Test 3 <u>0.078</u> | Test 6 <u>0.079</u> | Test 9 <u>0.079</u> |
| Average <u>0.078</u> | | |

J. Tank pressure, Level 3 Func D Sub G P: 716 psi Reg 700 psi

K. Remarks/Maintenance record: Annual inspection, replaced simulator return o-ring

24 Apr 2014
Date

24 APR 2014
Date

Deb Kashur
Field Inspector's Signature

Chadler E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:37

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:38 |
| 02 Std. Sol. | 0.022 | 10:39 |
| 03 Room Air | 0.000 | 10:39 |
| 04 Std. Sol. | 0.022 | 10:40 |
| 05 Room Air | 0.000 | 10:40 |
| 06 Std. Sol. | 0.022 | 10:41 |
| 07 Room Air | 0.000 | 10:42 |

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 29

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:32

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:32 |
| 02 Std. Sol. | 0.292 | 10:33 |
| 03 Room Air | 0.000 | 10:34 |
| 04 Std. Sol. | 0.295 | 10:34 |
| 05 Room Air | 0.000 | 10:35 |
| 06 Std. Sol. | 0.296 | 10:36 |
| 07 Room Air | 0.000 | 10:36 |

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 29

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:42

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 10:43 |
| 02 *Subject Test | INT* | 10:43 |
| 03 Room Air | 0.000 | 10:44 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 52
Expiration Date = 08/01/2014
County = 29 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:44

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 10:45 |
| 02 Room Air | RFI* | 10:45 |
| 03 Room Air | 0.000 | 10:46 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 52
Expiration Date = 08/01/2014
County = 29 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:46

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:47 |
| 02 Std. Gas | 0.078 | 10:47 |
| 03 Room Air | 0.000 | 10:47 |
| 04 Std. Gas | 0.078 | 10:48 |
| 05 Room Air | 0.000 | 10:48 |
| 06 Std. Gas | 0.078 | 10:49 |
| 07 Room Air | 0.000 | 10:49 |

Lot No = 16512080A2
Cyl No = 52
Exp Date = 08/01/2014
County = 29 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:49

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:50 |
| 02 Std. Gas | 0.079 | 10:50 |
| 03 Room Air | 0.000 | 10:51 |
| 04 Std. Gas | 0.078 | 10:51 |
| 05 Room Air | 0.000 | 10:52 |
| 06 Std. Gas | 0.079 | 10:52 |
| 07 Room Air | 0.000 | 10:52 |

Lot No = 16512080A2
Cyl No = 52
Exp Date = 08/01/2014
County = 29 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 ac

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = STAT 8164.13.00 06/09
04/24/2014 10:53

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:53 |
| 02 Std. Gas | 0.078 | 10:54 |
| 03 Room Air | 0.000 | 10:54 |
| 04 Std. Gas | 0.079 | 10:54 |
| 05 Room Air | 0.000 | 10:55 |
| 06 Std. Gas | 0.079 | 10:55 |
| 07 Room Air | 0.000 | 10:56 |

Lot No = 16512080A2
Cyl No = 52
Exp Date = 08/01/2014
County = 29 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080AC