

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003059

Location: TOXL

- | | Check When Done |
|---|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5114</u> Lot No. <u>20160314</u> Rep AC <u>0.010</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>MP3068</u> Lot No. <u>14102</u> Rep AC <u>0.400</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR5135</u> Lot No. <u>TC84</u> Solution <u>0.10 AC + 0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. <u>19615080A5</u>	Cyl No. <u>003</u>	Exp. Date <u>9/5/2017</u>
Test 1 <u>0.079</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

J. Tank pressure, Level 3 Func D Sub G P: 514 psi Reg ~525 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, REPLACED SIMULATOR

RETURN O-RING
4.21.2016
Date
25 Apr 16
Date

Charles E. Edr
Field Inspector's Signature
D. Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 3
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 04/07/2008
Slope 713
Intercept -692349

IR Calibration Date: 04/10/2015
 3um 9um

0th Coef(*100): -22715 -12336
1st Coef(*100): 274413 134965
2nd Coef(*100): 2480 1348
H2O adj (mg/l*10k): 437 307

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:36
02 Std. Sol.	0.009	13:37
03 Room Air	0.000	13:38
04 Std. Sol.	0.009	13:38
05 Room Air	0.000	13:39
06 Std. Sol.	0.009	13:40
07 Room Air	0.000	13:40

08 Sim Temp = 34.0°C

Simul Ser No = DR5114
Std Sol No = 201603H
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.010 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:42

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Sol.	0.395	13:43
03 Room Air	0.000	13:44
04 Std. Sol.	0.394	13:45
05 Room Air	0.000	13:45
06 Std. Sol.	0.395	13:46
07 Room Air	0.000	13:46

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:48

Test	AC	Time
01 Room Air	0.000	13:49
02 *Subject Test	INT*	13:49
03 Room Air	0.000	13:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 3
Expiration Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACE TONK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:51

Test	AC	Time
01 Diagnostic	OK	13:52
02 Room Air	0.000	13:52
03 *Subject Test	RFI*	13:52
04 Room Air	0.000	13:53

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 3
Expiration Date = 09/05/2017
County = 08 Oper No. = 666666

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.079	13:56
03 Room Air	0.000	13:56
04 Std. Gas	0.080	13:56
05 Room Air	0.000	13:57
06 Std. Gas	0.080	13:57
07 Room Air	0.000	13:58

Lot No = 19615080A5
Cyl No = 3
Exp Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 13:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:59
02 Std. Gas	0.080	13:59
03 Room Air	0.000	14:00
04 Std. Gas	0.079	14:00
05 Room Air	0.000	14:01
06 Std. Gas	0.081	14:01
07 Room Air	0.000	14:02

Lot No = 19615080A5
Cyl No = 3
Exp Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = TOXL 8164.13.00 06/09
04/21/2016 14:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Gas	0.080	14:03
03 Room Air	0.000	14:03
04 Std. Gas	0.080	14:04
05 Room Air	0.000	14:04
06 Std. Gas	0.080	14:05
07 Room Air	0.000	14:05

Lot No = 19615080A5
Cyl No = 3
Exp Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC