

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003071

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3066 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 14214080A6 Cyl No. 053 Exp. Date 7/5/2016

Test 1 <u>0.079</u>	Test 4 <u>0.078</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 430 psi Reg ~450 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, REPLACED

SIMULATOR RETURN O-RING

4.15.16
Date

25 Apr 16
Date

Chris E. Eder
Field Inspector's Signature

Db Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 15:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14214080A6
Standard Cyl #? 53
Standard Expiration? 07/05/2016
Oper No? 666666

Flow Cal. Date: 07/09/2010
Slope 806
Intercept -994737

IR Calibration Date: 04/10/2015
 3um 9um

0th Coef(*100): -11328 -10938
1st Coef(*100): 268087 134278
2nd Coef(*100): 3093 1368
H2O adj(mg/l*10k): 242 233

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 15:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:53
02 Std. Sol.	0.018	15:54
03 Room Air	0.000	15:55
04 Std. Sol.	0.018	15:55
05 Room Air	0.000	15:56
06 Std. Sol.	0.018	15:57
07 Room Air	0.000	15:57

08 Sim Temp = 34.0°C

Simul Ser No = MP3066
Std Sol No = 15160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low Acc AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 15:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:59
02 Std. Sol.	0.396	16:00
03 Room Air	0.000	16:00
04 Std. Sol.	0.397	16:01
05 Room Air	0.000	16:02
06 Std. Sol.	0.397	16:03
07 Room Air	0.000	16:03

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 16:05

Test	AC	Time
01 Room Air	0.000	16:06
02 *Subject Test	INT*	16:06
03 Room Air	0.000	16:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 14214080A6
Cyl No = 53
Expiration Date = 07/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 16:07

Test	AC	Time
01 Diagnostic	OK	16:09
02 Room Air	0.000	16:09
03 *Subject Test	RFI*	16:09
04 Room Air	0.000	16:10

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 14214080A6
Cyl No = 53
Expiration Date = 07/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 16:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:11
02 Std. Gas	0.079	16:11
03 Room Air	0.000	16:12
04 Std. Gas	0.078	16:12
05 Room Air	0.000	16:13
06 Std. Gas	0.079	16:13
07 Room Air	0.000	16:13

Lot No = 14214080A6
Cyl No = 53
Exp Date = 07/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 16:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:14
02 Std. Gas	0.078	16:15
03 Room Air	0.000	16:15
04 Std. Gas	0.078	16:15
05 Room Air	0.000	16:16
06 Std. Gas	0.079	16:16
07 Room Air	0.000	16:17

Lot No = 14214080A6
Cyl No = 53
Exp Date = 07/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TOXL 8164.13.00 06/09
04/15/2016 16:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:18
02 Std. Gas	0.079	16:18
03 Room Air	0.000	16:19
04 Std. Gas	0.079	16:19
05 Room Air	0.000	16:19
06 Std. Gas	0.079	16:20
07 Room Air	0.000	16:20

Lot No = 14214080A6
Cyl No = 53
Exp Date = 07/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000