

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004935

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3066 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 19615080A5 Cyl No. 003 Exp. Date 9/5/2017

Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.078</u>
Test 2 <u>0.079</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.078</u>	Test 6 <u>0.078</u>	Test 9 <u>0.078</u>

Average 0.078

J. Tank pressure, Level 3 Func D Sub G P: 414 psi Reg ~400 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, REPLACED SIMULATOR
RETURN O-RING, REPLACED BREATH HOSE O-RING

4.22.2016
Date
25 Apr 16
Date

Charles E. Egan
Field Inspector's Signature
Dch Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 3
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 07/12/2012
Slope 655
Intercept -510289

IR Calibration Date: 05/20/2015
 3um 9um

0th Coef(*100): -7871 -19684
1st Coef(*100): 264505 133108
2nd Coef(*100): 2865 1320
H2O adj(mg/l*10k): 200 298

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Sol.	0.019	10:07
03 Room Air	0.000	10:08
04 Std. Sol.	0.020	10:08
05 Room Air	0.000	10:09
06 Std. Sol.	0.019	10:09
07 Room Air	0.000	10:10

08 Sim Temp = 34.0°C

Simul Ser No = MP3066
Std Sol No = 15160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:12
02 Std. Sol.	0.399	10:13
03 Room Air	0.000	10:14
04 Std. Sol.	0.399	10:14
05 Room Air	0.000	10:15
06 Std. Sol.	0.399	10:16
07 Room Air	0.000	10:16

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
 0.400 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:21

Test	AC	Time
01 Room Air	0.000	10:22
02 *Subject Test	INT*	10:22
03 Room Air	0.000	10:23

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 3
Expiration Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:23

Test	AC	Time
01 Diagnostic	OK	10:24
02 Room Air	0.000	10:24
03 *Subject Test	RFI*	10:25
04 Room Air	0.000	10:25

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 3
Expiration Date = 09/05/2017
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK
0.08 CEE


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 09:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:57
02 Std. Gas	0.078	09:57
03 Room Air	0.000	09:58
04 Std. Gas	0.079	09:58
05 Room Air	0.000	09:59
06 Std. Gas	0.078	09:59
07 Room Air	0.000	09:59

Lot No = 19615080A5
Cyl No = 3
Exp Date = 09/05/2017
County = 08 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: *Calibration Check*
0.080Ac
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Gas	0.079	10:26
03 Room Air	0.000	10:27
04 Std. Gas	0.078	10:27
05 Room Air	0.000	10:28
06 Std. Gas	0.078	10:28
07 Room Air	0.000	10:28

Lot No = 19615080A5
Cyl No = 3
Exp Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = TOXL 8164.13.00 06/09
04/22/2016 10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:29
02 Std. Gas	0.078	10:29
03 Room Air	0.000	10:30
04 Std. Gas	0.078	10:30
05 Room Air	0.000	10:31
06 Std. Gas	0.078	10:31
07 Room Air	0.000	10:32

Lot No = 19615080A5
Cyl No = 3
Exp Date = 09/05/2017
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC