

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004948

Location: TOXL

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN MP3058 Lot No. 201501A Rep AC 0.015 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN MP3068 Lot No. 14102 Rep AC 0.400 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5135 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 14214080A6 Cyl No. 053 Exp. Date 7/5/2016

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 286 psi Reg ~300 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, REPLACED BATTERY

4.19.2016  
Date  
25 Apr 16  
Date

Charles E. Ede  
Field Inspector's Signature  
Deb Shenaver  
Reviewed by

Intoxilyzer Test Record and Checklist  
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer	Alcohol Analyzer
North Dakota Model 8000	SN 80-004948
Location = TOXL	8164.13.00 06/09
04/19/2016	15:53

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-|=|  
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|  
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:	Enabled
Start Test Sequence:	DABACABA
Display Prelim Rslt?	Yes
Display Third Digit?	Yes
Inhib Printer(Y/N)?	No
Display Volume?	No
Disable On Memfull?	Yes
# of Print Copies?	1
Select Std (D/W/I)?	Dry
Standard Value?	0.080
Standard Lot #?	14214080A6
Standard Cyl #?	53
Standard Expiration?	07/05/2016
Oper No?	666666

Flow Cal. Date:	07/14/2010
Slope	778
Intercept	-795769

IR Calibration Date:	07/14/2010
	3um                      9um

-----		
0th Coef(*100):	-31061	-21991
1st Coef(*100):	265097	133596
2nd Coef(*100):	802	1052
H2O adj(mg/l*10k):	675	417

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
 CHARLES EDER

Remarks:

*PRINT TEST*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.13.00 06/09  
04/19/2016      15:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:55
02 Std. Sol.	0.013	15:56
03 Room Air	0.000	15:56
04 Std. Sol.	0.013	15:57
05 Room Air	0.000	15:57
06 Std. Sol.	0.013	15:58
07 Room Air	0.000	15:58

08 Sim Temp = 34.0°C

Simul Ser No = MP3058  
Std Sol No = 201501A  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.13.00 06/09  
04/19/2016      16:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Sol.	0.397	16:06
03 Room Air	0.000	16:07
04 Std. Sol.	0.395	16:07
05 Room Air	0.000	16:08
06 Std. Sol.	0.395	16:09
07 Room Air	0.000	16:09

08 Sim Temp = 34.0°C

Simul Ser No = MP3068  
Std Sol No = 14102  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: HIGH AC  
0.400 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004948  
Location = TOXL                    8164.13.00 06/09  
04/19/2016                            16:11

Test	AC	Time
01 Room Air	0.000	16:12
02 *Subject Test	INT*	16:13
03 Room Air	0.000	16:13

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 14214080A6  
Cyl No = 53  
Expiration Date = 07/05/2016  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Interferent Check*  
*0.10AC + 0.05% ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.13.00 06/09  
04/19/2016      16:13

Test	AC	Time
01 Diagnostic	OK	16:14
02 Room Air	0.000	16:15
03 Reference	RFI*	16:15
04 Room Air	0.000	16:15

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 14214080A6  
Cyl No = 53  
Expiration Date = 07/05/2016  
County = 08      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI DETECT*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.13.00 06/09  
04/19/2016      16:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:18
02 Std. Gas	0.081	16:18
03 Room Air	0.000	16:18
04 Std. Gas	0.081	16:19
05 Room Air	0.000	16:19
06 Std. Gas	0.081	16:20
07 Room Air	0.000	16:20

Lot No = 14214080A6  
Cyl No = 53  
Exp Date = 07/05/2016  
County = 08      Oper No. = 666666

  
Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.13.00 06/09  
04/19/2016      16:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:21
02 Std. Gas	0.081	16:21
03 Room Air	0.000	16:22
04 Std. Gas	0.080	16:22
05 Room Air	0.000	16:23
06 Std. Gas	0.081	16:23
07 Room Air	0.000	16:23

Lot No = 14214080A6  
Cyl No = 53  
Exp Date = 07/05/2016  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

CALIBRATION CHECK  
0.080 AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004948  
Location = TOXL      8164.13.00 06/09  
04/19/2016      16:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:24
02 Std. Gas	0.081	16:24
03 Room Air	0.000	16:25
04 Std. Gas	0.081	16:25
05 Room Air	0.000	16:26
06 Std. Gas	0.081	16:26
07 Room Air	0.000	16:27

Lot No = 14214080A6  
Cyl No = 53  
Exp Date = 07/05/2016  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000