

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005947

Location: Bismarck PD

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>15160</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR5190</u> Lot No. <u>I3S4</u> Solution <u>0.10 AC + 0.25% acetone</u>
<u>DS</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. <u>29614080A2</u>	Cyl No. <u>27</u>	Exp. Date <u>12/05/2016</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.078</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

J. Tank pressure, Level 3 Func D Sub G P: 245 psi Reg 255 psi

K. Remarks/Maintenance record: Annual inspection, replaced breath hose O-ring.

26 Apr 2016
Date
4-26-2016
Date

Deb Shanaver
Field Inspector's Signature
Charles E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
04/26/2016 13:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:53
02 Std. Sol.	0.020	13:54
03 Room Air	0.000	13:54
04 Std. Sol.	0.019	13:55
05 Room Air	0.000	13:56
06 Std. Sol.	0.020	13:56
07 Room Air	0.000	13:57

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 15160

County = 08

Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
04/26/2016 13:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Sol.	0.305	13:48
03 Room Air	0.000	13:49
04 Std. Sol.	0.306	13:50
05 Room Air	0.000	13:50
06 Std. Sol.	0.306	13:51
07 Room Air	0.000	13:52

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

e^{DS} High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
04/26/2016 13:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Gas	0.079	13:36
03 Room Air	0.000	13:36
04 Std. Gas	0.079	13:36
05 Room Air	0.000	13:37
06 Std. Gas	0.079	13:37
07 Room Air	0.000	13:38

Lot No = 29614080A2
Cyl No = 27
Exp Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
04/26/2016 13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Gas	0.079	13:39
03 Room Air	0.000	13:40
04 Std. Gas	0.079	13:40
05 Room Air	0.000	13:40
06 Std. Gas	0.079	13:41
07 Room Air	0.000	13:41

Lot No = 29614080A2
Cyl No = 27
Exp Date = 12/05/2016
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = BSPD 8164.13.00 06/09
04/26/2016 13:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:42
02 Std. Gas	0.079	13:42
03 Room Air	0.000	13:43
04 Std. Gas	0.078	13:43
05 Room Air	0.000	13:44
06 Std. Gas	0.079	13:44
07 Room Air	0.000	13:45

Lot No = 29614080A2

Cyl No = 27

Exp Date = 12/05/2016

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Form 106-I8000