

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Phil Wolf

Location: Cavalier, ND

Intoxilyzer Serial No.: 80-002669

Lot Number: 659357

Cylinder Number: 9

Expiration Date: 4/3/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.077 %

Test 2: 0.077 %

Test 3: 0.077 %

(ATTACH TEST RECORD)

P. Wolf
Chemical Test Operator Signature

4/27/10
Date

Deb Kashum
Reviewed By

10 May 2010
Date

80-002669
p. 81 of 82
JAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
04/27/2010 21:16

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 21:16 |
| 02 Std. Gas | 0.077 | 21:17 |
| 03 Room Air | 0.000 | 21:17 |
| 04 Std. Gas | 0.077 | 21:17 |
| 05 Room Air | 0.000 | 21:18 |
| 06 Std. Gas | 0.077 | 21:18 |
| 07 Room Air | 0.000 | 21:19 |


Lot No = 659358

Cyl No = 34

Exp Date = 09/03/2011

County = 34

Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
P. 82482
JFM