

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003057 Location: TOXL

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 10290 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 11110 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. Rep AC 0.100 ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080A1 Cyl No. 49 Exp. Date 10/01/2013
Test 1 0.081 Test 4 0.081 Test 7 0.082
Test 2 0.082 Test 5 0.080 Test 8 0.080
Test 3 0.081 Test 6 0.082 Test 9 0.081
Average 0.081

L. Tank pressure, Level 3 Func D Sub G P: 708 psi Reg 700 psi

M. Remarks/Maintenance record: Annual inspection

27 Apr 2012
Date

Deb Kashur
Field Inspector's Signature

01 MAY 2012
Date

Charles E. Egan
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 09:28

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 49
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 11/29/2010
Slope 653
Intercept -334316

IR Calibration Date: 06/14/2011
 3um 9um

0th Coef(*100): -6449 357
1st Coef(*100): 246728 129732
2nd Coef(*100): 5256 2318
H2O adj(mg/l*10k): 576 529

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 10:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Sol.	0.020	10:04
03 Room Air	0.000	10:05
04 Std. Sol.	0.021	10:05
05 Room Air	0.000	10:06
06 Std. Sol.	0.021	10:07
07 Room Air	0.000	10:07

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 10:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Sol.	0.301	10:24
03 Room Air	0.000	10:24
04 Std. Sol.	0.300	10:25
05 Room Air	0.000	10:25
06 Std. Sol.	0.300	10:26
07 Room Air	0.000	10:27

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 13:33

Test	AC	Time
01 Room Air	0.000	13:33
02 *Subject Test	INT*	13:34
03 Room Air	0.000	13:34

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 49
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 13:47

Test	AC	Time
01 Diagnostic	OK	13:48
02 Room Air	RFI*	13:48
03 Room Air	0.000	13:49

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 49
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 09:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:30
02 Std. Gas	0.081	09:30
03 Room Air	0.000	09:30
04 Std. Gas	0.082	09:31
05 Room Air	0.000	09:31
06 Std. Gas	0.081	09:32
07 Room Air	0.000	09:32

Lot No = 23411080A1

Cyl No = 49

Exp Date = 10/01/2013

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 09:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:33
02 Std. Gas	0.081	09:34
03 Room Air	0.000	09:34
04 Std. Gas	0.080	09:35
05 Room Air	0.000	09:35
06 Std. Gas	0.082	09:36
07 Room Air	0.000	09:36

Lot No = 23411080A1

Cyl No = 49

Exp Date = 10/01/2013

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = TOXL 8164.13.00 06/09
04/27/2012 09:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:56
02 Std. Gas	0.082	09:56
03 Room Air	0.000	09:57
04 Std. Gas	0.080	09:57
05 Room Air	0.000	09:58
06 Std. Gas	0.081	09:58
07 Room Air	0.000	09:59

Lot No = 23411080A1
Cyl No = 49
Exp Date = 10/01/2013
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC