



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Shannon Liebold</u>		
Location: <u>Mercer County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: 80-003061 ^{DK} <u>80-003060</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>43</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Shannon P. Liebold
 Chemical Test Operator Signature

4-28-11
 Date

Deb Kashner
 Reviewed By

06 May 2011
 Date

80-003060
 P. 102 of 103
 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
04/28/2011 08:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:49
02 Std. Gas	0.080	08:50
03 Room Air	0.000	08:50
04 Std. Gas	0.080	08:51
05 Room Air	0.000	08:51
06 Std. Gas	0.080	08:52
07 Room Air	0.000	08:52


Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060
P.103 of 103
vsm