

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-604952

Location: HETT

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR515 Lot No. 12111 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR543 Lot No. 13080 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. ICS2 Rep AC 0.10 AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 165/2080A2 Cyl No. 03 Exp. Date 08/01/2014

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.078</u>	Test 6 <u>0.080</u>	Test 9 <u>0.079</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 980 psi Reg 1000 psi

K. Remarks/Maintenance record: Annual inspection

25 Apr 2014
Date

28 APR 14
Date

Deb Kashner
Field Inspector's Signature

Charles E. Eden
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:25

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 3
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 01/19/2010
Slope 694
Intercept -638616

IR Calibration Date: 01/19/2010
3um 9um

	3um	9um
0th Coef(*100):	-32832	-34603
1st Coef(*100):	262566	131011
2nd Coef(*100):	1418	975
H2O adj(mg/l*10k):	686	550

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:27
02 Std. Sol.	0.017	12:27
03 Room Air	0.000	12:28
04 Std. Sol.	0.016	12:29
05 Room Air	0.000	12:29
06 Std. Sol.	0.017	12:30
07 Room Air	0.000	12:30

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 01

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:32
02 Std. Sol.	0.287	12:33
03 Room Air	0.000	12:34
04 Std. Sol.	0.288	12:34
05 Room Air	0.000	12:35
06 Std. Sol.	0.289	12:36
07 Room Air	0.000	12:36

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 01

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

High AC ~~0.0300 A~~ 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:37

Test	AC	Time
01 Room Air	0.000	12:38
02 *Subject Test	INT*	12:39
03 Room Air	0.000	12:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 03
Expiration Date = 08/01/2014
County = 01 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10 AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:40

Test	AC	Time
01 Diagnostic	OK	12:41
02 Room Air	RFI*	12:41
03 Room Air	0.000	12:42

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 03
Expiration Date = 08/01/2014
County = 01 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:44
02 Std. Gas	0.079	12:44
03 Room Air	0.000	12:45
04 Std. Gas	0.078	12:45
05 Room Air	0.000	12:46
06 Std. Gas	0.078	12:46
07 Room Air	0.000	12:46

Lot No = 16512080A2
Cyl No = 03
Exp Date = 08/01/2014
County = 01 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:48
02 Std. Gas	0.079	12:48
03 Room Air	0.000	12:49
04 Std. Gas	0.079	12:49
05 Room Air	0.000	12:50
06 Std. Gas	0.080	12:50
07 Room Air	0.000	12:51

Lot No = 16512080A2
Cyl No = 03
Exp Date = 08/01/2014
County = 01 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = HETT 8164.13.00 06/09
04/25/2014 12:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:51
02 Std. Gas	0.079	12:52
03 Room Air	0.000	12:52
04 Std. Gas	0.080	12:53
05 Room Air	0.000	12:53
06 Std. Gas	0.079	12:53
07 Room Air	0.000	12:54

Lot No = 16512080A2
Cyl No = 03
Exp Date = 08/01/2014
County = 01 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000