

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004188

Location: BOWM (BOWMAN Co. S. O.)

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5132 Lot No. ICS #2 Solution 0.10 AC + 0.05% ACE ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 14214080A6 Cyl No. 67 Exp. Date 7/5/16  
Test 1 0.081 Test 4 0.081 Test 7 0.081  
Test 2 0.082 Test 5 0.081 Test 8 0.081  
Test 3 0.082 Test 6 0.082 Test 9 0.082  
Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 1017 psi Reg 1000+ psi

K. Remarks/Maintenance record: REPLACED SIMULATOR RETURN 0-12ING,

ANNUAL INSPECTION

4.22.15  
Date

28 Apr 2015 ✓  
Date

Charles E. Eder  
Field Inspector's Signature

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-004188  
Location = BOWM                   8164.13.00 06/09  
04/22/2015                           13:01

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?            No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?            0.080  
Standard Lot #?            14214080A6  
Standard Cyl #?            067  
Standard Expiration?       07/05/2016  
Oper No?                    666666

Flow Cal. Date:            07/05/2011  
Slope                       668  
Intercept                  -597204

IR Calibration Date:       04/05/2012  
                            3um                   9um

	3um	9um
0th Coef(*100):	-15090	-16406
1st Coef(*100):	263383	126966
2nd Coef(*100):	6533	2371
H2O adj(mg/l*10k):	801	809

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004188  
Location = BOWM                    8164.13.00 06/09  
04/22/2015                            13:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:18
02 Std. Sol.	0.020	13:19
03 Room Air	0.000	13:20
04 Std. Sol.	0.020	13:20
05 Room Air	0.000	13:21
06 Std. Sol.	0.020	13:22
07 Room Air	0.000	13:22


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 06

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*  
*0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
04/22/2015      13:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Sol.	0.302	13:30
03 Room Air	0.000	13:30
04 Std. Sol.	0.298	13:31
05 Room Air	0.000	13:32
06 Std. Sol.	0.295	13:32
07 Room Air	0.000	13:33

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 06

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC  
0.300 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004188  
Location = BOWM                8164.13.00 06/09  
04/22/2015                        13:14

Test	AC	Time
01 Room Air	0.000	13:15
02 *Subject Test	INT*	13:15
03 Room Air	0.000	13:16

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 14214080A6  
Cyl No = 067  
Expiration Date = 07/05/2016  
County = 06                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent Check*  
*0.10AC + 0.05% ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004188  
Location = BOWM                8164.13.00 06/09  
04/22/2015                        13:02

Test	AC	Time
01 Diagnostic	OK	13:04
02 Room Air	0.000	13:05
03 *Subject Test	RFI*	13:05
04 Room Air	0.000	13:06

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 14214080A6  
Cyl No = 067  
Expiration Date = 07/05/2016  
County = 06                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004188  
Location = BOWM                8164.13.00 06/09  
04/22/2015                        13:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:07
02 Std. Gas	0.081	13:07
03 Room Air	0.000	13:08
04 Std. Gas	0.082	13:08
05 Room Air	0.000	13:09
06 Std. Gas	0.082	13:09
07 Room Air	0.000	13:10

Lot No = 14214080A6  
Cyl No = 067  
Exp Date = 07/05/2016  
County = 06                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004188  
Location = BOWM                    8164.13.00 06/09  
04/22/2015                            13:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:10
02 Std. Gas	0.081	13:11
03 Room Air	0.000	13:11
04 Std. Gas	0.081	13:11
05 Room Air	0.000	13:12
06 Std. Gas	0.082	13:12
07 Room Air	0.000	13:13

Lot No = 14214080A6  
Cyl No = 067  
Exp Date = 07/05/2016  
County = 06                            Oper No. = 666666

  
Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = BOWM      8164.13.00 06/09  
04/22/2015      13:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:36
02 Std. Gas	0.081	13:36
03 Room Air	0.000	13:37
04 Std. Gas	0.081	13:37
05 Room Air	0.000	13:38
06 Std. Gas	0.082	13:38
07 Room Air	0.000	13:39

Lot No = 14214080A6  
Cyl No = 67  
Exp Date = 07/05/2016  
County = 06      Oper No. = 666666

  
Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC