



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Steven R Fischer</u>		
Location: <u>Ward County Jail, Mint</u>		
Intoxilyzer® Serial Number: <u>80-003066</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>32</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Steven R Fischer
Chemical Test Operator Signature

4-29-11
Date

Deb Kashner
Reviewed By

06 May 2011
Date

80-003066
p. 138 of 139
SEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = WARD 8164.13.00 06/09
04/29/2011 17:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:27
02 Std. Gas	0.080	17:27
03 Room Air	0.000	17:28
04 Std. Gas	0.080	17:28
05 Room Air	0.000	17:28
06 Std. Gas	0.079	17:29
07 Room Air	0.000	17:29

Lot No = 659358
Cyl No = 32
Exp Date = 09/03/2011
County = 51 Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-003066
p. 139 of 139
FM