

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David J. Zeis

Location: Cavalier County Sheriff's Office (Langdon)

Intoxilyzer Serial No.: 80-004188

Lot Number: 659358

Cylinder Number: 53

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 077 %

Test 2: 0. 077 %

Test 3: 0. 077 %

(ATTACH TEST RECORD)

David J. Zeis  
Chemical Test Operator Signature

04/30/10  
Date

Deb Kashner  
Reviewed By

10 May 2010  
Date

80-004188  
P. 74/79 KM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = LANG      8164.13.00 06/09  
04/30/2010      10:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Gas	0.077	10:06
03 Room Air	0.000	10:06
04 Std. Gas	0.077	10:07
05 Room Air	0.000	10:07
06 Std. Gas	0.077	10:08
07 Room Air	0.000	10:08

Lot No = 659358  
Cyl No = 53  
Exp Date = 09/03/2011  
County = 10      Oper No. = 067518

  
\_\_\_\_\_  
Operator Signature  
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-004188  
P. 79/79 KM