



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Ray Weber</u>		
Location: <u>Trail Co. S.D., Hillsboro, N.D.</u>		
Intoxilyzer® Serial Number: <u>80-004949</u>		
Gas Lot Number: <u>1540080A1</u>	Gas Cylinder Number: <u>415</u>	Expiration Date: <u>07/01/2012</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>078</u> % Test 2: 0. <u>078</u> % Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Ray Weber
 Chemical Test Operator Signature

4/30/2011
 Date

Deb Kashur
 Reviewed By

16 May 2011
 Date

80-004949
 P. 100 of 101
 RM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = HILL 8164.13.00 06/09
04/30/2011 13:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:41
02 Std. Gas	0.078	13:41
03 Room Air	0.000	13:42
04 Std. Gas	0.078	13:42
05 Room Air	0.000	13:42
06 Std. Gas	0.079	13:43
07 Room Air	0.000	13:43

Lot No = 15410080A1

Cyl No = 15

Exp Date = 07/01/2012

County = 49

Oper No. = 058405

Ray Weber

Operator Signature

RAY G. WEBER

Remarks:

Form 106-I8000

80-004949
P. 101 of 101
JTM