

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004203

Location: Rugby - Heart of Amer. Corr.

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR5132 Lot No. 10290 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR3803 Lot No. 11110 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR3846 Lot No. — Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. ~~23411080A1~~ 23411080A1<sup>DK</sup> Cyl No. 54 Exp. Date 10/01/2013  
Test 1 0.080 Test 4 0.081 Test 7 0.080  
Test 2 0.080 Test 5 0.080 Test 8 0.080  
Test 3 0.080 Test 6 0.081 Test 9 0.080  
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 1012 psi Reg 1000 psi

M. Remarks/Maintenance record: Annual inspection

30 Apr 2012  
Date

Deb Keshur  
Field Inspector's Signature

01 MAY 12  
Date

Charles E. Eder  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      14:26

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 23411080A1  
Standard Cyl #? 54  
Standard Expiration? 10/01/2013  
Oper No? 777777

Flow Cal. Date: 01/28/2010  
Slope 719  
Intercept -711903

IR Calibration Date: 01/28/2010  
3um 9um

-----  
0th Coef(\*100): -17469 -28710  
1st Coef(\*100): 264216 135022  
2nd Coef(\*100): 2778 1081  
H2O adj(mg/l\*10k): 430 605

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      15:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:07
02 Std. Sol.	0.019	15:08
03 Room Air	0.000	15:08
04 Std. Sol.	0.019	15:09
05 Room Air	0.000	15:09
06 Std. Sol.	0.019	15:10
07 Room Air	0.000	15:11

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 35

Oper No. = 77777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      15:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:01
02 Std. Sol.	0.300	15:02
03 Room Air	0.000	15:03
04 Std. Sol.	0.297	15:03
05 Room Air	0.000	15:04
06 Std. Sol.	0.296	15:05
07 Room Air	0.000	15:05

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 35

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      15:13

Test	AC	Time
01 Room Air	0.000	15:14
02 *Subject Test	INT*	15:14
03 Room Air	0.000	15:15

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 54  
Expiration Date = 10/01/2013  
County = 35      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      15:15

Test	AC	Time
01 Diagnostic	OK	15:16
02 Room Air	RFI*	15:16
03 Room Air	0.000	15:17

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 54  
Expiration Date = 10/01/2013  
County = 35      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.080	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.080	14:38
05 Room Air	0.000	14:38
06 Std. Gas	0.080	14:39
07 Room Air	0.000	14:39

Lot No = 23411080A1  
Cyl No = 54  
Exp Date = 10/01/2013  
County = 35      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC  
Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004203  
Location = RUGB                    8164.13.00 06/09  
04/30/2012                            14:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.081	14:41
03 Room Air	0.000	14:41
04 Std. Gas	0.080	14:42
05 Room Air	0.000	14:42
06 Std. Gas	0.081	14:42
07 Room Air	0.000	14:43

Lot No = 23411080A1  
Cyl No = 54  
Exp Date = 10/01/2013  
County = 35                            Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = RUGB      8164.13.00 06/09  
04/30/2012      14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Gas	0.080	14:44
03 Room Air	0.000	14:45
04 Std. Gas	0.080	14:45
05 Room Air	0.000	14:46
06 Std. Gas	0.080	14:46
07 Room Air	0.000	14:46

Lot No = 23411080A1  
Cyl No = 54  
Exp Date = 10/01/2013  
County = 35      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC