

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004947

Location: Renville Co SO

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5192</u> Lot No. <u>10290</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3803</u> Lot No. <u>11110</u> Rep AC <u>0.300</u>                               | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>11611080A2</u>	Cyl No. <u>143</u>	Exp. Date <u>06/01/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 721 psi Reg 720 psi

M. Remarks/Maintenance record: Annual inspection.

30 Apr 2012  
Date

Dob Kashur  
Field Inspector's Signature

01 MAY 12  
Date

Charles E. Egan  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = MOHL      8164.13.00 06/09  
04/30/2012      10:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:51
02 Std. Sol.	0.016	10:52
03 Room Air	0.000	10:52
04 Std. Sol.	0.017	10:53
05 Room Air	0.000	10:54
06 Std. Sol.	0.017	10:54
07 Room Air	0.000	10:55

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 38

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004947  
Location = MOHL                8164.13.00 06/09  
04/30/2012                        11:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Sol.	0.286	11:01
03 Room Air	0.000	11:02
04 Std. Sol.	0.287	11:02
05 Room Air	0.000	11:03
06 Std. Sol.	0.288	11:04
07 Room Air	0.000	11:04

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 38

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004947  
Location = MOHL                8164.13.00 06/09  
04/30/2012                        11:06

Test	AC	Time
01 Room Air	0.000	11:07
02 *Subject Test	INT*	11:08
03 Room Air	0.000	11:08

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 11611080A2  
Cyl No = 143  
Expiration Date = 06/01/2013  
County = 38                        Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004947  
Location = MOHL                8164.13.00 06/09  
04/30/2012                        11:10

Test	AC	Time
01 Diagnostic	OK	11:11
02 Room Air	RFI*	11:11
03 Room Air	0.000	11:11

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 11611080A2  
Cyl No = 143  
Expiration Date = 06/01/2013  
County = 38                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = MOHL      8164.13.00 06/09  
04/30/2012      10:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:27
02 Std. Gas	0.079	10:27
03 Room Air	0.000	10:28
04 Std. Gas	0.079	10:28
05 Room Air	0.000	10:28
06 Std. Gas	0.079	10:29
07 Room Air	0.000	10:29

Lot No = 11611080A2  
Cyl No = 143  
Exp Date = 06/01/2013  
County = ~~08~~ *38 DK*      Oper No. = 777777

*Deb Kasher*  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004947  
Location = MOHL                8164.13.00 06/09  
04/30/2012                        10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.079	10:31
03 Room Air	0.000	10:31
04 Std. Gas	0.080	10:31
05 Room Air	0.000	10:32
06 Std. Gas	0.079	10:32
07 Room Air	0.000	10:33

Lot No = 11611080A2  
Cyl No = 143  
Exp Date = 06/01/2013  
County = 38                      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
 North Dakota Model 8000      SN 80-004947  
 Location = MOHL      8164.13.00 06/09  
 04/30/2012      10:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:35
02 Std. Gas	0.079	10:35
03 Room Air	0.000	10:35
04 Std. Gas	0.079	10:36
05 Room Air	0.000	10:36
06 Std. Gas	0.080	10:37
07 Room Air	0.000	10:37

Lot No = 11611080A2  
 Cyl No = 143  
 Exp Date = 06/01/2013  
 County = 38      Oper No. = 777777

Deb Kasher  
 Operator Signature  
 DEB KASHUR

Remarks:

*Calibration check 0.080 AC*