

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004950

Location: Harvey PD

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 10290 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 11110 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. — Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080A1 Cyl No. 22 Exp. Date 10/01/2013
Test 1 0.079 Test 4 0.079 Test 7 0.078
Test 2 0.079 Test 5 0.079 Test 8 0.079
Test 3 0.078 Test 6 0.079 Test 9 0.078
Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 973 psi Reg 980 psi

M. Remarks/Maintenance record: Annual inspection

30 Apr 2012
Date

Deh Kashur
Field Inspector's Signature

01 MAY 12
Date

[Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
04/30/2012 16:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:50
02 Std. Sol.	0.020	16:51
03 Room Air	0.000	16:52
04 Std. Sol.	0.021	16:52
05 Room Air	0.000	16:53
06 Std. Sol.	0.021	16:53
07 Room Air	0.000	16:54

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 52

Oper No. = 777777

Deb Keshur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
04/30/2012 16:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Sol.	0.298	16:56
03 Room Air	0.000	16:56
04 Std. Sol.	0.299	16:57
05 Room Air	0.000	16:58
06 Std. Sol.	0.299	16:58
07 Room Air	0.000	16:59

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 52

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
04/30/2012 17:01

Test	AC	Time
01 Room Air	0.000	17:02
02 *Subject Test	INT*	17:02
03 Room Air	0.000	17:03

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 22
Expiration Date = 10/01/2013
County = 52 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 AC + acetone
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
04/30/2012 16:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:39
02 Std. Gas	0.079	16:39
03 Room Air	0.000	16:40
04 Std. Gas	0.079	16:40
05 Room Air	0.000	16:40
06 Std. Gas	0.078	16:41
07 Room Air	0.000	16:41

Lot No = 23411080A1
Cyl No = 22
Exp Date = 10/01/2013
County = 52 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
04/30/2012 16:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:43
02 Std. Gas	0.079	16:43
03 Room Air	0.000	16:44
04 Std. Gas	0.079	16:44
05 Room Air	0.000	16:45
06 Std. Gas	0.079	16:45
07 Room Air	0.000	16:46

Lot No = 23411080A1
Cyl No = 22
Exp Date = 10/01/2013
County = 52 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = HARV 8164.13.00 06/09
04/30/2012 16:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:46
02 Std. Gas	0.078	16:47
03 Room Air	0.000	16:47
04 Std. Gas	0.079	16:47
05 Room Air	0.000	16:48
06 Std. Gas	0.078	16:48
07 Room Air	0.000	16:49

Lot No = 23411080A1

Cyl No = 22

Exp Date = 10/01/2013

County = 52

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000