

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004193

Location: Fort Yates PD

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use < 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS 2 Rep AC 0.10 AC + 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 25 Exp. Date 08/01/2014

Test 1 <u>0.079</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 705 psi Reg 700 psi

K. Remarks/Maintenance record: annual inspection

30 Apr 14
Date
30 APR 14
Date

Deb Kasher
Field Inspector's Signature
Charles Elder
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:57

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled

Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 25
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 06/04/2009
Slope 692
Intercept -611138

IR Calibration Date: 04/08/2013
 3um 9um

0th Coef(*100): -16195 -13956
1st Coef(*100): 263840 129613
2nd Coef(*100): 2879 1369
H2O adj(mg/l*10k): 539 519

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Sol.	0.021	11:56
03 Room Air	0.000	11:57
04 Std. Sol.	0.022	11:58
05 Room Air	0.000	11:58
06 Std. Sol.	0.022	11:59
07 Room Air	0.000	11:59

08 Sim Temp = 34.0°C

Simul Ser No = DR5115
Std Sol No = 12111
County = 43 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:51
02 Std. Sol.	0.300	11:51
03 Room Air	0.000	11:52
04 Std. Sol.	0.303	11:53
05 Room Air	0.000	11:53
06 Std. Sol.	0.303	11:54
07 Room Air	0.000	11:54

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 43

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:30

Test	AC	Time
01 Room Air	0.000	11:31
02 *Subject Test	INT*	11:31
03 Room Air	0.000	11:32

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 25
Expiration Date = 08/01/2014
County = 43 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:33

Test	AC	Time
01 Diagnostic	.. OK	11:33
02 Room Air	RFI*	11:33
03 Room Air	0.000	11:34

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 25
Expiration Date = 08/01/2014
County = 43 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:37
02 Std. Gas	0.079	11:37
03 Room Air	0.000	11:38
04 Std. Gas	0.079	11:38
05 Room Air	0.000	11:39
06 Std. Gas	0.080	11:39
07 Room Air	0.000	11:40

Lot No = 16512080A2..
Cyl No = 25
Exp Date = 08/01/2014
County = 43 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Gas	0.080	11:41
03 Room Air	0.000	11:42
04 Std. Gas	0.079	11:42
05 Room Air	0.000	11:43
06 Std. Gas	0.079	11:43
07 Room Air	0.000	11:43

Lot No = 16512080A2
Cyl No = 25
Exp Date = 08/01/2014
County = 43 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = FYPD 8164.13.00 06/09
04/30/2014 11:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:45
02 Std. Gas	0.079	11:45
03 Room Air	0.000	11:46
04 Std. Gas	0.080	11:46
05 Room Air	0.000	11:47
06 Std. Gas	0.080	11:47
07 Room Air	0.000	11:48

Lot No = 16512080A2
Cyl No = 25
Exp Date = 08/01/2014
County = 43 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000