

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson I 51

Intoxilyzer Serial No.: 80 003066

Lot Number: DMW  
23 659358

Cylinder Number: DMW  
9-3-11 23

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082%

Test 2: 0. 082%

Test 3: 0. 083%

(ATTACH TEST RECORD)

David M Wallace  
Chemical Test Operator Signature

5-1-10  
Date

Deb Kaizer  
Reviewed By

10 May 2010  
Date


Intoxilyzer Test Record and Checklist  
NDCAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = DKPD      8164.13.00 06/09  
05/01/2010      09:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Gas	0.082	09:22
03 Room Air	0.000	09:23
04 Std. Gas	0.082	09:23
05 Room Air	0.000	09:23
06 Std. Gas	0.083	09:24
07 Room Air	0.000	09:24

Lot No = 659358  
Cyl No = 23  
Exp Date = 09/03/2011  
County = 45      Oper No. = 068332



Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-18000

80-003066  
P. 85 of 85  
FM