



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|--|-------------------------------|------------------------------------|
| Chemical Test Operator Name: <u>Justin Wellenich</u> | | |
| Location: <u>Cooperstown</u> | | |
| Intoxilyzer® Serial Number: <u>80-004203</u> | | |
| Gas Lot Number: <u>0591108041</u> | Gas Cylinder Number: <u>5</u> | Expiration Date: <u>02/01/2013</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed | | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) | | |
| Test 1: 0.79 % <u>0.079</u> DK | | |
| Test 2: 0.79 % <u>0.079</u> | | |
| Test 3: 0.80 % <u>0.080</u> | | |
| (ATTACH TEST RECORD) | | |

[Signature]
Chemical Test Operator Signature

05-01-2011
Date

Deb Kashner
Reviewed By

06 May 2011
Date

80-004203
p. 103 of 1084
JTW


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = COOP 8164.13.00 06/09
05/01/2011 13:56

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 13:57 |
| 02 Std. Gas | 0.079 | 13:58 |
| 03 Room Air | 0.000 | 13:58 |
| 04 Std. Gas | 0.079 | 13:59 |
| 05 Room Air | 0.000 | 13:59 |
| 06 Std. Gas | 0.080 | 13:59 |
| 07 Room Air | 0.000 | 14:00 |

Lot No = 05911080A1
Cyl No = 5
Exp Date = 02/01/2013
County = 20 Oper No. = 011017


Operator Signature
JUSTIN A. WALLERICH

Remarks:

Form 106-I8000

80-004203
R104 of 104
FM