



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Sector/Breath Alcohol Program
SFN #9282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|--|--------------------------------|--------------------------------|
| Chemical Test Operator Name: <u>Patrick Hudson</u> | | |
| Location: <u>Ward So</u> | | |
| Intoxilyzer® Serial Number: <u>80-023058</u> | | |
| Gas Lot Number: <u>659358</u> | Gas Cylinder Number: <u>32</u> | Expiration Date: <u>9/3/11</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed | | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) | | |
| Test 1: 0. <u>080</u> % | | |
| Test 2: 0. <u>080</u> % | | |
| Test 3: 0. <u>082</u> % | | |
| (ATTACH TEST RECORD) | | |

Patrick Hudson
Chemical Test Operator Signature

5/2/10
Date

Deb Kashner
Reviewed By

17 May 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = WARD 8164.13.00 06/09
05/02/2010 19:47

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 19:48 |
| 02 Std. Gas | 0.080 | 19:49 |
| 03 Room Air | 0.000 | 19:49 |
| 04 Std. Gas | 0.080 | 19:49 |
| 05 Room Air | 0.000 | 19:50 |
| 06 Std. Gas | 0.082 | 19:50 |
| 07 Room Air | 0.000 | 19:51 |


Lot No = 659358

Cyl No = 32

Exp Date = 09/03/2011

County = 51

Oper No. = 049509



Operator Signature
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003058
p. 83-83
JCH