

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002666

Location: Dickinson PD

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5132</u> Lot No. <u>10290</u> Rep AC <u>0.070</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3863</u> Lot No. <u>1110</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3846</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080 A1</u>	Cyl No. <u>21</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.086</u>	Test 5 <u>0.080</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 674 psi Reg 680 psi

M. Remarks/Maintenance record: Annual inspection

02 May 2012
Date

Deb Kashner
Field Inspector's Signature

02 MAY 2012
Date

Charles E. Edr
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 21
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 09/10/2007
Slope 677
Intercept -603196

IR Calibration Date: 09/10/2007
 3um 9um

0th Coef(*100): -14148 -8312
1st Coef(*100): 274152 135499
2nd Coef(*100): 1387 855
H2O adj(mg/l*10k): 430 237

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:42
02 Std. Sol.	0.020	12:42
03 Room Air	0.000	12:43
04 Std. Sol.	0.021	12:44
05 Room Air	0.000	12:44
06 Std. Sol.	0.020	12:45
07 Room Air	0.000	12:45

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 45

Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Low Ac ^{BK} ~~0.021~~ 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:47
02 Std. Sol.	0.296	12:48
03 Room Air	0.000	12:48
04 Std. Sol.	0.295	12:49
05 Room Air	0.000	12:49
06 Std. Sol.	0.294	12:50
07 Room Air	0.000	12:51

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 45

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:52

Test	AC	Time
01 Room Air	0.000	12:53
02 *Subject Test	INT*	12:53
03 Room Air	0.000	12:54

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 21
Expiration Date = 10/01/2013
County = 45 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:54

Test	AC	Time
01 Diagnostic	OK	12:55
02 Room Air	RFI*	12:55
03 Room Air	0.000	12:55

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 21
Expiration Date = 10/01/2013
County = 45 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:22
02 Std. Gas	0.080	12:23
03 Room Air	0.000	12:23
04 Std. Gas	0.080	12:24
05 Room Air	0.000	12:24
06 Std. Gas	0.079	12:25
07 Room Air	0.000	12:25

Lot No = 23411080A1
Cyl No = 21
Exp Date = 10/01/2013
County = 45 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:29
02 Std. Gas	0.080	12:29
03 Room Air	0.000	12:30
04 Std. Gas	0.080	12:30
05 Room Air	0.000	12:30
06 Std. Gas	0.080	12:31
07 Room Air	0.000	12:31

Lot No = 23411080A1
Cyl No = 21
Exp Date = 10/01/2013
County = 45 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = DKPD 8164.13.00 06/09
05/02/2012 12:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:35
02 Std. Gas	0.080	12:35
03 Room Air	0.000	12:36
04 Std. Gas	0.079	12:36
05 Room Air	0.000	12:36
06 Std. Gas	0.080	12:37
07 Room Air	0.000	12:37

Lot No = 23411080A1
Cyl No = 21
Exp Date = 10/01/2013
County = 45 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC