

**INTOXILYZER 8000 INSPECTION**

Serial No.: 80-004196

Location: McLean Co - Washburn

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>    </u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5115</u> Lot No. <u>    </u> Rep AC <u>0.298</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>659357</u>	Cyl No. <u>47</u>	Exp. Date <u>09/03/2011</u>
Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G . P: 764 psi Reg 150 psi

M. Remarks/Maintenance record: Annual inspection

03 May 2010  
Date

Deb Kashur  
Field Inspector's Signature

15 May 2010  
Date

Mary Pearson  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:24
02 Std. Sol.	0.019	11:25
03 Room Air	0.000	11:26
04 Std. Sol.	0.019	11:26
05 Room Air	0.000	11:27
06 Std. Sol.	0.020	11:28
07 Room Air	0.000	11:28

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 28

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.021AC*

Form 106-I8000

*80-004196  
P. 82/88 JAM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Sol.	0.292	11:30
03 Room Air	0.000	11:31
04 Std. Sol.	0.293	11:31
05 Room Air	0.000	11:32
06 Std. Sol.	0.292	11:33
07 Room Air	0.000	11:33

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 28

Oper No. = 777777

*Deb Kasher*

Operator Signature

DEB KASHUR

Remarks:

*Linearity test 0.298 AC*

Form 106-I8000

80-004196  
P. 8/3/88 *VFH*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:21

Test	AC	Time
01 Room Air	0.000	11:22
02 *Subject Test	INT*	11:22
03 Room Air	0.000	11:23

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659357  
Cyl No = 47  
Expiration Date = 09/03/2011  
County = 28      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check*

*0.100 AC  
+ acetone*

Form 106-I8000

*80-004196  
P. 89/88 PM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:00

Test	AC	Time
01 Diagnostic	OK	11:01
02 Room Air	RFI*	11:01
03 Room Air	0.000	11:02

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659357  
Cyl No = 47  
Expiration Date = 09/03/2011  
County = 28      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Keshur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004196  
P. 85/88 YH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:04
02 Std. Gas	0.081	11:04
03 Room Air	0.000	11:05
04 Std. Gas	0.081	11:05
05 Room Air	0.000	11:06
06 Std. Gas	0.081	11:06
07 Room Air	0.000	11:06

Lot No = 659357  
Cyl No = 47  
Exp Date = 09/03/2011  
County = 28      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check      0.080 AC*

Form 106-I8000

*80-004196  
P. 8/1/88 KM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Gas	0.082	11:08
03 Room Air	0.000	11:08
04 Std. Gas	0.081	11:09
05 Room Air	0.000	11:09
06 Std. Gas	0.081	11:09
07 Room Air	0.000	11:10

Lot No = 659357

Cyl No = 47

Exp Date = 09/03/2011

County = 28

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004196  
P. 87/88 KA



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
05/03/2010      11:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Gas	0.081	11:12
03 Room Air	0.000	11:12
04 Std. Gas	0.081	11:13
05 Room Air	0.000	11:13
06 Std. Gas	0.082	11:13
07 Room Air	0.000	11:14

Lot No = 659357

Cyl No = 47

Exp Date = 09/03/2011

County = 28

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004196  
P. 08/08 JEM