



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: Adam Krabbenhoft
Location: La Moure County Courthouse
Intoxilyzer® Serial Number: 80-004942
Gas Lot Number: 709352 Gas Cylinder Number: 14 Expiration Date: 12-23-11
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0.080% Test 2: 0.080% Test 3: 0.080% <p style="text-align: right;">(ATTACH TEST RECORD)</p>

  
\_\_\_\_\_  
Chemical Test Operator Signature

5-3-11  
\_\_\_\_\_  
Date

Deb Keshner  
\_\_\_\_\_  
Reviewed By

06 May 2011  
\_\_\_\_\_  
Date

80-004942  
P. 65 of 66  
vfm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004942  
Location = LAMO      8164.13.00 06/09  
05/03/2011      10:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:37
02 Std. Gas	0.080	10:37
03 Room Air	0.000	10:38
04 Std. Gas	0.080	10:38
05 Room Air	0.000	10:39
06 Std. Gas	0.080	10:39
07 Room Air	0.000	10:39

Lot No = 709352  
Cyl No = 14  
Exp Date = 12/23/2011  
County = 23      Oper No. = 020828



Operator Signature  
ADAM R. KRABBENHOFT

Remarks:

ACA Test

Form 106-I8000

80-004942  
P. 26 of 66  
RM