

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004189

Location: WILL (WILLIAMS C.L.E.C.)

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u>                                     | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>1110</u> Rep AC <u>0.300</u>                             | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3377</u> Lot No. <u>—</u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>23411080A1</u>	Cyl No. <u>23</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 642 psi Reg 650 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

03 MAY 2012  
Date

Charles E. Egan  
Field Inspector's Signature

07 May 12  
Date

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:05

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?


Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            23  
Standard Expiration?       10/01/2013  
Oper No?                    666666

Flow Cal. Date:            02/02/2011  
Slope                        628  
Intercept                  -452608

IR Calibration Date:        02/02/2011  
                                  3um                    9um  
-----  
0th Coef(\*100):            -22075                -26801  
1st Coef(\*100):            257433                132480  
2nd Coef(\*100):            2132                   1061  
H2O adj(mg/l\*10k):        437                    430

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004189  
Location = WILL      8164.13.00 06/09  
05/03/2012      16:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:21
02 Std. Sol.	0.017	16:22
03 Room Air	0.000	16:23
04 Std. Sol.	0.018	16:23
05 Room Air	0.000	16:24
06 Std. Sol.	0.017	16:25
07 Room Air	0.000	16:25

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 53

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:27
02 Std. Sol.	0.298	16:28
03 Room Air	0.000	16:28
04 Std. Sol.	0.298	16:29
05 Room Air	0.000	16:30
06 Std. Sol.	0.298	16:30
07 Room Air	0.000	16:31

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 53

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

LINEARITY TEST  
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:32

Test	AC	Time
01 Room Air	0.000	16:34
02 *Subject Test	INT*	16:34
03 Room Air	0.000	16:35

\*Invalid Test  
Interferent Detected

Sub Name = SPIRIT, DISCOVER THE  
Sub DOB = 10/22/1969  
Sub Sex = Male                            Weight = 150  
Test = DUI                            Cit = QWERTYUIOPLKJHG  
Dr. Lic. = ND/SPI691234  
Simul Ser No = 000000  
Std Sol No = 0000  
County = 53                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT CHECK*  
*0.100 AC + ACETONE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:36

Test	AC	Time
01 Diagnostic	OK	16:37
02 Room Air	RFI*	16:38
03 Room Air	0.000	16:38

\*Invalid Test  
Inhibited - RFI

Sub Name = SPIRIT, DISCOVER THE  
Sub DOB = 10/22/1969  
Sub Sex = Male                            Weight = 150  
Test = DUI                            Cit = ZXCVBNMASDFGHJK  
Dr. Lic. = ND/SPI691234  
Simul Ser No = 000000  
Std Sol No = 0000  
County = 53                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:41
02 Std. Gas	0.079	16:41
03 Room Air	0.000	16:42
04 Std. Gas	0.079	16:42
05 Room Air	0.000	16:43
06 Std. Gas	0.080	16:43
07 Room Air	0.000	16:44

Lot No = 23411080A1  
Cyl No = 23  
Exp Date = 10/01/2013  
County = 53                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *CALIBRATION CHECK*  
*0.080 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:45
02 Std. Gas	0.079	16:45
03 Room Air	0.000	16:46
04 Std. Gas	0.079	16:46
05 Room Air	0.000	16:47
06 Std. Gas	0.079	16:47
07 Room Air	0.000	16:47

Lot No = 23411080A1  
Cyl No = 23  
Exp Date = 10/01/2013  
County = 53                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004189  
Location = WILL                    8164.13.00 06/09  
05/03/2012                            16:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:48
02 Std. Gas	0.079	16:49
03 Room Air	0.000	16:49
04 Std. Gas	0.079	16:50
05 Room Air	0.000	16:50
06 Std. Gas	0.079	16:51
07 Room Air	0.000	16:51

Lot No = 23411080A1  
Cyl No = 23  
Exp Date = 10/01/2013  
County = 53                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC

Form 106-I8000