

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004200

Location: CROS (DIVIDE Co. SO.)

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>13</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.078</u>	Test 4 <u>0.080</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.078</u>
Test 3 <u>0.078</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 476 psi Reg 500 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

<u>03 MAY 2012</u> Date	<u>Charles E. Edr</u> Field Inspector's Signature
<u>07 May 2012</u> Date	<u>Deb Kashner</u> Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 13:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Sol.	0.018	13:55
03 Room Air	0.000	13:56
04 Std. Sol.	0.018	13:57
05 Room Air	0.000	13:57
06 Std. Sol.	0.019	13:58
07 Room Air	0.000	13:58


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 12

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 14:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:02
02 Std. Sol.	0.288	14:02
03 Room Air	0.000	14:03
04 Std. Sol.	0.290	14:04
05 Room Air	0.000	14:04
06 Std. Sol.	0.291	14:05
07 Room Air	0.000	14:06


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 12

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 14:07

Test	AC	Time
01 Room Air	0.000	14:08
02 *Subject Test	INT*	14:08
03 Room Air	0.000	14:09

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = ZXCVBNMLKJHGFD
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100AC + ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 14:11

Test	AC	Time
01 Diagnostic	OK	14:12
02 Room Air	RFI*	14:12
03 Room Air	0.000	14:12

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 12 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 14:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:15
02 Std. Gas	0.078	14:16
03 Room Air	0.000	14:16
04 Std. Gas	0.078	14:16
05 Room Air	0.000	14:17
06 Std. Gas	0.078	14:17
07 Room Air	0.000	14:18

Lot No = 23411080A1
Cyl No = 13
Exp Date = 10/01/2013
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 14:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:19
02 Std. Gas	0.080	14:19
03 Room Air	0.000	14:20
04 Std. Gas	0.079	14:20
05 Room Air	0.000	14:21
06 Std. Gas	0.079	14:21
07 Room Air	0.000	14:22

Lot No = 23411080A1
Cyl No = 13
Exp Date = 10/01/2013
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
05/03/2012 14:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Gas	0.079	14:23
03 Room Air	0.000	14:23
04 Std. Gas	0.078	14:24
05 Room Air	0.000	14:24
06 Std. Gas	0.078	14:24
07 Room Air	0.000	14:25

Lot No = 23411080A1
Cyl No = 13
Exp Date = 10/01/2013
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080AC