

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004940

Location: BURK (BURKE CO. S.O.)

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u>---</u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>31</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.078</u>	Test 4 <u>0.077</u>	Test 7 <u>0.078</u>
Test 2 <u>0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.077</u>	Test 6 <u>0.077</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 936 psi Reg 925 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

<u>03 MAY 2012</u> Date	<u>Charles E. Egan</u> Field Inspector's Signature
<u>07 May 12</u> Date	<u>Dub Kashner</u> Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 11:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 31
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 720
Intercept -746294

IR Calibration Date: 06/22/2010
 3um 9um

0th Coef(*100):	-10417	-10800
1st Coef(*100):	259947	132173
2nd Coef(*100):	2872	1619
H2O adj(mg/l*10k):	604	506

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 11:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Sol.	0.017	11:42
03 Room Air	0.000	11:43
04 Std. Sol.	0.018	11:43
05 Room Air	0.000	11:44
06 Std. Sol.	0.018	11:45
07 Room Air	0.000	11:45

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 0000
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 11:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Sol.	0.295	11:50
03 Room Air	0.000	11:51
04 Std. Sol.	0.297	11:51
05 Room Air	0.000	11:52
06 Std. Sol.	0.299	11:53
07 Room Air	0.000	11:53

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 07

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LINEARITY TEST

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 11:55

Test	AC	Time
01 Room Air	0.000	11:56
02 *Subject Test	INT*	11:56
03 Room Air	0.000	11:57

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT CHECK
0.100AC + ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 12:04

Test	AC	Time
01 Diagnostic	OK	12:05
02 Room Air	RFI*	12:05
03 Room Air	0.000	12:06

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOP
Dr. Lic. = ND/DIS821456
Simul Ser No = 000000
Std Sol No = 0000
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 12:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Gas	0.078	12:09
03 Room Air	0.000	12:09
04 Std. Gas	0.078	12:10
05 Room Air	0.000	12:10
06 Std. Gas	0.077	12:11
07 Room Air	0.000	12:11

Lot No = 23411080A1
Cyl No = 31
Exp Date = 10/01/2013
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 12:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:12
02 Std. Gas	0.077	12:12
03 Room Air	0.000	12:13
04 Std. Gas	0.078	12:13
05 Room Air	0.000	12:14
06 Std. Gas	0.077	12:14
07 Room Air	0.000	12:15

Lot No = 23411080A1
Cyl No = 31
Exp Date = 10/01/2013
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = BURK 8164.13.00 06/09
05/03/2012 12:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:16
02 Std. Gas	0.078	12:16
03 Room Air	0.000	12:17
04 Std. Gas	0.078	12:17
05 Room Air	0.000	12:17
06 Std. Gas	0.078	12:18
07 Room Air	0.000	12:18

Lot No = 23411080A1
Cyl No = 31
Exp Date = 10/01/2013
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000