

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002666

Location: HARV

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>ICS#1</u> Rep AC <u>0.10 AC + 0.05% ACETONE</u> ^{OCEE} | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>22</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 1071 psi Reg 1100 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

03 MAY 13
Date
03 May 13
Date

Charles E. Ed
Field Inspector's Signature
Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 22
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 12/31/2012
Slope 657
Intercept -418473

IR Calibration Date: 12/31/2012
 3um 9um

0th Coef(*100): -31967 -16114
1st Coef(*100): 279454 136158
2nd Coef(*100): 1837 1190
H2O adj(mg/l*10k): 527 293

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Sol.	0.017	11:37
03 Room Air	0.000	11:38
04 Std. Sol.	0.017	11:38
05 Room Air	0.000	11:39
06 Std. Sol.	0.017	11:40
07 Room Air	0.000	11:40

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 52

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC (Sim. Sol.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:55
02 Std. Sol.	0.294	11:56
03 Room Air	0.000	11:56
04 Std. Sol.	0.297	11:57
05 Room Air	0.000	11:58
06 Std. Sol.	0.298	11:58
07 Room Air	0.000	11:59

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 52

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LINEARITY TEST
0.300 AC (SIM. SOL.)

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:32

Test	AC	Time
01 Room Air	0.000	11:33
02 *Subject Test	INT*	11:33
03 Room Air	0.000	11:34

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH1234
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 22
Expiration Date = 10/01/2013
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100% AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:29

Test	AC	Time
01 Diagnostic	OK	11:30
02 Room Air	RFI*	11:30
03 Room Air	0.000	11:31

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 22
Expiration Date = 10/01/2013
County = 52 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Gas	0.080	11:43
03 Room Air	0.000	11:43
04 Std. Gas	0.079	11:43
05 Room Air	0.000	11:44
06 Std. Gas	0.079	11:44
07 Room Air	0.000	11:45

Lot No = 23411080A1
Cyl No = 22
Exp Date = 10/01/2013
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:46
02 Std. Gas	0.079	11:46
03 Room Air	0.000	11:47
04 Std. Gas	0.080	11:47
05 Room Air	0.000	11:47
06 Std. Gas	0.080	11:48
07 Room Air	0.000	11:48


Lot No = 23411080A1

Cyl No = 22

Exp Date = 10/01/2013

County = 52

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/03/2013 11:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:49
02 Std. Gas	0.079	11:50
03 Room Air	0.000	11:50
04 Std. Gas	0.080	11:50
05 Room Air	0.000	11:51
06 Std. Gas	0.080	11:51
07 Room Air	0.000	11:52

Lot No = 23411080A1
Cyl No = 22
Exp Date = 10/01/2013
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)