

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000
~~80-004950~~ CEE Serial Number: 80-004950

Agency being assigned instrument: OFFICE OF ATTORNEY GENERAL
CRIME LAB

Date 5/3/2016

Transferred from:

Name (print) Amanda Mosser

Signature Amanda Mosser

Agency Tioga Police Department

Transferred to:

Name (print) CHARLES E. EDER

Signature Charles Eder

Agency O.A.G. - CRIME LAB

Reason for transfer/remarks:

- INSTRUMENT STILL WITHIN SPECIFICATION ($\pm 5\%$).
- HIGH AC RUNNING 4-5% HIGH
- CALIBRATION ADJUSTMENT WILL BE PERFORMED

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____