



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Justin Wallach</u>		
Location: <u>Coopers town</u>		
Intoxilyzer® Serial Number: <u>80-003071</u>		
Gas Lot Number: <u>05911080A1</u>	Gas Cylinder Number: <u>5</u>	Expiration Date: <u>02-01-2013</u>
<input checked="" type="checkbox"/> New Cylinder Installed <input type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>79</u> % 0.079 DK Test 2: 0. <u>79</u> % 0.079 Test 3: 0. <u>79</u> % 0.079		

(ATTACH TEST RECORD)

[Signature]  
 Chemical Test Operator Signature

05/04/11  
 Date

Deb Kasher  
 Reviewed By

06 May 2011  
 Date

80-003071  
 P 131 of 132  
 SFM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = COOP      8164.13.00 06/09  
05/04/2011      14:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:14
02 Std. Gas	0.079	14:15
03 Room Air	0.000	14:15
04 Std. Gas	0.079	14:16
05 Room Air	0.000	14:16
06 Std. Gas	0.079	14:16
07 Room Air	0.000	14:17

Lot No = 05911080A1  
Cyl No = 5  
Exp Date = 02/01/2013  
County = 20      Oper No. = 011017

  
\_\_\_\_\_  
Operator Signature  
JUSTIN A. WALLERICH

Remarks:

Form 106-I8000

80-003071  
p. 132 of 132  
KM