



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Joseph Knowski</i>		
Location <i>Devils lake LEC</i>	Intoxilyzer® Serial Number <i>80-003061</i>	
Gas Lot Number <i>05911080A1</i>	Gas Cylinder Number <i>1</i>	Gas Expiration Date <i>02-01-2013</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 080 AC
 - B. 0. 080 AC
 - C. 0. 079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Joseph Knowski</i>	Date <i>5-4-2012</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>14 May 2012</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = DEVL 8164.13.00 06/09
05/04/2012 14:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Gas	0.080	14:23
03 Room Air	0.000	14:23
04 Std. Gas	0.080	14:24
05 Room Air	0.000	14:24
06 Std. Gas	0.079	14:24
07 Room Air	0.000	14:25


Lot No = 05911080A1

Cyl No = 1

Exp Date = 02/01/2013

County = 36

Oper No. = 109208


Operator Signature
JOSEPH J. KNOWSKI

Remarks: