

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004953 Location: NEW T (TAT PD)

- |   | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes?  | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?   | <u>✓</u>        |
| C. Is the breath tube heated?   | <u>✓</u>        |
| D. Is the diagnostic check complete?  | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| F. Print test. (Attach test record.)  | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3455</u> Lot No. <u>10290</u> Rep AC <u>0.020</u>                                     | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>11110</u> Rep AC <u>0.300</u>                            | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3377</u> Lot No. <u>—</u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. <u>2341108041</u>	Cyl No. <u>3</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.078</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u><del>0.06</del> 0.078</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.078</u>	Test 6 <u>0.079</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 955 psi Reg 950 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

<u>04 MAY 12</u> Date	<u>Charles E. Edin</u> Field Inspector's Signature
<u>07 May 12</u> Date	<u>Deb Kashner</u> Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-004953  
Location = NEWT               8164.13.00 06/09  
05/04/2012                               13:58

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       No  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?           23411080A1  
Standard Cyl #?           03  
Standard Expiration?       10/01/2013  
Oper No?                   666666

Flow Cal. Date:           01/19/2010  
Slope                      691  
Intercept                 -628068

IR Calibration Date:       06/08/2010  
                            3um                   9um

-----  
0th Coef(\*100):       -10343               -18245  
1st Coef(\*100):       249276               131289  
2nd Coef(\*100):       4014                   1439  
H2O adj(mg/l\*10k):   697                   655

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:    PRINT TEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = NEWT      8164.13.00 06/09  
05/04/2012      14:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:11
02 Std. Sol.	0.018	14:12
03 Room Air	0.000	14:12
04 Std. Sol.	0.018	14:13
05 Room Air	0.000	14:14
06 Std. Sol.	0.018	14:14
07 Room Air	0.000	14:15

08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = ~~31~~ CFE  
60

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = NEWT      8164.13.00 06/09  
05/04/2012      14:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:23
02 Std. Sol.	0.303	14:24
03 Room Air	0.000	14:24
04 Std. Sol.	0.303	14:25
05 Room Air	0.000	14:26
06 Std. Sol.	0.303	14:26
07 Room Air	0.000	14:27


08 Sim Temp = 34.0°C

Simul Ser No = 000000

Std Sol No = 0000

County = 60

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*LINEARITY TEST*  
*0.300 AC*






Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = NEWT      8164.13.00 06/09  
05/04/2012      14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.078	14:37
03 Room Air	0.000	14:37
04 Std. Gas	0.078	14:38
05 Room Air	0.000	14:38
06 Std. Gas	0.078	14:39
07 Room Air	0.000	14:39

Lot No = 23411080A1  
Cyl No = 3  
Exp Date = 10/01/2013  
County = 60      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004953  
Location = NEWT            8164.13.00 06/09  
05/04/2012                            14:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.079	14:40
03 Room Air	0.000	14:41
04 Std. Gas	0.078	14:41
05 Room Air	0.000	14:42
06 Std. Gas	0.079	14:42
07 Room Air	0.000	14:43

Lot No = 23411080A1  
Cyl No = 3  
Exp Date = 10/01/2013  
County = 60                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = NEWT      8164.13.00 06/09  
05/04/2012      14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:43
02 Std. Gas	0.079	14:44
03 Room Air	0.000	14:44
04 Std. Gas	0.079	14:44
05 Room Air	0.000	14:45
06 Std. Gas	0.078	14:45
07 Room Air	0.000	14:46

Lot No = 23411080A1  
Cyl No = 3  
Exp Date = 10/01/2013  
County = 60      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC