

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-002666

Location: HARV (HARVEY P.D.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5132 Lot No. ICS#2 Solution 0.10AC + 0.05% ACE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 3391300044 Cyl No. 43 Exp. Date 1/1/2016
Test 1 0.079 Test 4 0.082 Test 7 0.081
Test 2 0.080 Test 5 0.081 Test 8 0.081
Test 3 0.080 Test 6 0.081 Test 9 0.081
Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 547 psi Reg 550 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5.4.15
Date

04 May 15
Date

[Signature]
Field Inspector's Signature

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 13:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 43
Standard Expiration? 01/01/2016
Oper No? 666666

Flow Cal. Date: 12/31/2012
Slope 657
Intercept -418473

IR Calibration Date: 12/31/2012

	3um	9um
0th Coef(*100):	-31967	-16114
1st Coef(*100):	279454	136158
2nd Coef(*100):	1837	1190
H2O adj(mg/l*10k):	527	293

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 13:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Sol.	0.018	13:52
03 Room Air	0.000	13:52
04 Std. Sol.	0.018	13:53
05 Room Air	0.000	13:54
06 Std. Sol.	0.018	13:54
07 Room Air	0.000	13:55

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 201501B
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 14:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Sol.	0.295	14:04
03 Room Air	0.000	14:04
04 Std. Sol.	0.296	14:05
05 Room Air	0.000	14:06
06 Std. Sol.	0.295	14:06
07 Room Air	0.000	14:07

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 14270
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 13:38

Test	AC	Time
01 Room Air	0.000	13:39
02 *Subject Test	INT*	13:40
03 Room Air	0.000	13:40

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 33913080A4
Cyl No = 43
Expiration Date = 01/01/2016
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 13:41

Test	AC	Time
01 Diagnostic	OK	13:42
02 Room Air	0.000	13:42
03 *Subject Test	RFI*	13:42
04 Room Air	0.000	13:43

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 33913080A4
Cyl No = 43
Expiration Date = 01/01/2016
County = 52 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 13:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:46
02 Std. Gas	0.079	13:46
03 Room Air	0.000	13:47
04 Std. Gas	0.080	13:47
05 Room Air	0.000	13:48
06 Std. Gas	0.080	13:48
07 Room Air	0.000	13:49

Lot No = 33913080A4
Cyl No = 43
Exp Date = 01/01/2016
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 14:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:13
02 Std. Gas	0.082	14:13
03 Room Air	0.000	14:14
04 Std. Gas	0.081	14:14
05 Room Air	0.000	14:15
06 Std. Gas	0.081	14:15
07 Room Air	0.000	14:16

Lot No = 33913080A4
Cyl No = 43
Exp Date = 01/01/2016
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
05/04/2015 14:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:16
02 Std. Gas	0.081	14:17
03 Room Air	0.000	14:17
04 Std. Gas	0.081	14:18
05 Room Air	0.000	14:18
06 Std. Gas	0.081	14:18
07 Room Air	0.000	14:19

Lot No = 33913080A4
Cyl No = 43
Exp Date = 01/01/2016
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC